



Annalise Critical Care AI

Performance Guide

Product	Release
Annalise Critical Care AI	3.10 which includes: <ul style="list-style-type: none">Annalise Viewer version 3.7Annalise Backend version 3.8Annalise Integration Adapter version 3.9



ANNALISE-AI PTY LTD
Level P
24 Campbell St
Sydney NSW 2000
AUSTRALIA
www.annalise.ai
support@annalise.ai



Table of Contents

Introduction	1
Guide details	2
Overview	2
Symbol glossary	2
Standalone performance evaluation	3
Background	4
Overview	4
Ground truth interpretations	4
Subgroup analysis	5
Additional information	5
Detection accuracy as Area Under the Curve (AUROC)	6
Overall sensitivity and specificity	7
Patient age	10
Patient sex	14
Patient race and ethnicity	18
Equipment manufacturer	34
Series slice thickness for CTB findings	51
Presence of co-existing findings or abnormalities	53
Other	67
Triage effectiveness performance	71
Triage effectiveness	72
Definition	72
Assessing triage effectiveness	72
Results	72
Support	73
Support	74
Support and feedback	74

Introduction

Guide details

Overview

This document outlines the performance of the Annalise Critical Care AI device.






The performance results contained in this document are based on the dataset used by Annalise.ai to evaluate the device.

Differences in demographics, imaging equipment, image quality or other variables may result in changes in performance.

For general user information refer to the *Annalise Critical Care AI User Guide*.

Symbol glossary

Definitions of symbols that may appear on the device or in the related documentation are listed below.

Symbol	Information
	Manufacturer
	Medical device
	Prescription only
	Indicates a warning or caution
	Read the instructions for use

Standalone performance evaluation

Background

Overview

A standalone performance evaluation of the device was performed on a test dataset of cases obtained within the United States (chest X-ray cases were obtained from four sites and brain CT cases were obtained from five sites).

Ground truth interpretations

The reference standard ('ground truth') for the test dataset was determined by two ground truthers, with a third ground truth radiologist used in the event of disagreement. All ground truthers were US board-certified radiologists with a thoracic sub-specialty for chest x-ray evaluations and a neuroradiology sub-specialty for CT brain evaluations.

The device performed inference on the test dataset and the results were compared with the ground truth interpretations to evaluate the standalone performance of the device for each finding:

Chest X-ray (CXR)	CT brain (CTB)
<ul style="list-style-type: none">• pleural effusion• pneumoperitoneum• pneumothorax• tension pneumothorax• vertebral compression fracture	<ul style="list-style-type: none">• acute subarachnoid hemorrhage• acute subdural/epidural hematoma• intra-axial hemorrhage• intraventricular hemorrhage• mass effect• obstructive hydrocephalus*• vasogenic edema

*The device does not differentiate between acute, subacute or chronic.

Performance was measured in terms of detection accuracy as Area Under the Receiver-Operating Characteristic Curve (AUROC), and sensitivity and specificity.

The results of these standalone performance evaluations are summarized in the following pages.



During the obstructive hydrocephalus standalone performance evaluation which had a total test dataset of size of n=175 for slice thickness $\leq 1.5\text{mm}$ and n=191 for slice thickness $> 1.5\text{mm}$ to $\leq 5.0\text{mm}$, a subset of cases (n=4 for slice thickness $\leq 1.5\text{mm}$ and n=6 for slice thickness $> 1.5\text{mm}$ and $\leq 5.0\text{mm}$) were incorrectly filtered out by the device prior to clinical AI model inference.

As a result, no performance result for these cases could be obtained and the cases were excluded from the following analyses.

Subgroup analysis

The test dataset included a range of patient demographics and imaging equipment manufacturers. Subgroup analysis was performed for the following:

- patient age
- patient sex
- patient ethnicity (distribution)
- patient race (including distribution)
- equipment manufacturer
- series thickness for CTB findings
- presence of co-existing findings or abnormalities (including distribution)
- presence of specific co-existing findings or abnormalities (intra-axial hemorrhage only)
- acute subdural/epidural hematoma location
- presence of mimics for obstructive hydrocephalus

Detection accuracy as sensitivity and specificity for each sub-group is summarized in each of the following tables.

Additional information

The following additional information is helpful for interpreting results.

Finding/s	Additional information
Pleural effusion	<ul style="list-style-type: none">• specificity may be reduced in the presence of scarring and/or pleural thickening• standalone performance evaluation was performed on a dataset that included supine and erect positioning• use of this device with prone positioning may result in differences in performance
Pneumoperitoneum	<ul style="list-style-type: none">• standalone performance evaluation was performed on a dataset that included supine and erect positioning where most cases were of unilateral right-sided and bilateral pneumoperitoneum• use of this device with prone positioning and for unilateral left-sided pneumoperitoneum may result in differences in performance
Vertebral compression fracture	<ul style="list-style-type: none">• standalone performance evaluation was performed on a dataset that included only erect positioning• use of this device with supine positioning may result in differences in performance

Detection accuracy as Area Under the Curve (AUROC)

Results

The AUROC and distribution of ground truth positive/negative cases is outlined below.



For CTB findings, AUROC was evaluated at two slice thickness ranges for each finding.

CXR: Detection accuracy as Area Under the Curve (AUROC)

Finding	Positive cases	Negative cases	AUROC (95% CI)
Pleural effusion	482	507	0.977 (0.969-0.984)
Pneumoperitoneum	105	190	0.987 (0.976-0.994)
Pneumothorax	418	542	0.984 (0.976-0.990)
Tension pneumothorax	125	835	0.989 (0.984-0.994)
Vertebral compression fracture	272	318	0.972 (0.960-0.982)

CTB: Detection accuracy as Area Under the Curve (AUROC)

Finding	Slice thickness range	Positive cases	Negative cases	AUROC (95%)
Acute subarachnoid hemorrhage	≤1.5mm	147 (63.4%)	85 (36.6%)	0.993 (0.985-0.998)
	>1.5mm & ≤5.0mm	182 (63.4%)	105 (36.6%)	0.967 (0.946-0.983)
Acute subdural/epidural hematoma	≤1.5mm	303 (75.6%)	98 (24.4%)	0.972 (0.957-0.984)
	>1.5mm & ≤5.0mm	397 (74.6%)	135 (25.4%)	0.942 (0.921-0.959)
Intra-axial hemorrhage	≤1.5mm	480 (68.4%)	222 (31.6%)	0.968 (0.954-0.978)
	>1.5mm & ≤5.0mm	587 (67.1%)	288 (32.9%)	0.965 (0.952-0.976)
Intraventricular hemorrhage	≤1.5mm	73 (48.7%)	77 (51.3%)	0.986 (0.968-0.997)
	>1.5mm & ≤5.0mm	91 (49.5%)	93 (50.5%)	0.983 (0.966-0.994)
Mass effect	≤1.5mm	493 (78.8%)	133 (21.2%)	0.987 (0.979-0.993)
	>1.5mm & ≤5.0mm	594 (78.0%)	168 (22.0%)	0.983 (0.974-0.991)
Obstructive hydrocephalus	≤1.5mm	75 (42.9%)	100 (57.1%)	0.988 (0.972-0.998)
	>1.5mm & ≤5.0mm	84 (44.0%)	107 (56.0%)	0.987 (0.969-0.997)
Vasogenic edema	≤1.5mm	60 (35.9%)	107 (64.1%)	0.981 (0.961-0.993)
	>1.5mm & ≤5.0mm	99 (33.8%)	194 (66.2%)	0.988 (0.977-0.996)

Overall sensitivity and specificity

Results

Sensitivity and specificity were evaluated at a range of operating points for each finding. For CTB, these were evaluated at two slice thickness ranges.

An operating point configuration is provided as part of the organization's initial installation. An organization may select a different operating point from the range of points validated during the device's standalone performance evaluation.

Different operating points may be selected for each finding for an organization based on the following considerations:

- prevalence of disease in the population
- significance of false positives and false negatives, and
- criticality of the clinical condition.

Changes in configuration to apply these operating points can only be made in consultation with Annalise.ai.

Sensitivity and specificity results are shown as follows:

- CXR: at each operating point
- CTB: at each operating point and each slice thickness range

See:

- [CXR: Sensitivity and specificity by operating point, page 8](#)
- [CTB: Sensitivity and specificity by operating point, page 9](#)

CXR: Sensitivity and specificity by operating point

Finding	Operating point	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion	0.380	96.7 (95.0,98.1)	86.8 (83.6,89.5)
	0.425	94.4 (92.3,96.5)	89.5 (86.8,92.1)
	0.450	92.9 (90.7,95.0)	91.3 (88.6,93.7)
	0.475	89.8 (87.1,92.3)	93.7 (91.5,95.9)
	0.500	87.6 (84.6,90.5)	95.5 (93.5,97.0)
Pneumoperitoneum	0.250	96.2 (92.4,99.0)	87.9 (83.2,92.1)
	0.300	94.3 (89.5,98.1)	90.5 (86.3,94.2)
	0.350	92.4 (86.7,97.1)	93.7 (90.0,96.8)
	0.400	91.4 (85.7,96.2)	95.8 (92.6,98.4)
	0.450	87.6 (81.0,93.3)	98.4 (96.3,100.0)
Pneumothorax	0.200	97.1 (95.5,98.6)	88.2 (85.4,90.8)
	0.250	96.2 (94.3,98.1)	91.9 (89.5,94.1)
	0.300	95.0 (92.8,97.1)	94.1 (91.9,95.9)
	0.350	93.1 (90.7,95.5)	95.6 (93.7,97.2)
	0.400	90.7 (88.0,93.3)	96.7 (95.0,98.2)
Tension pneumothorax	0.225	96.0 (92.0,99.2)	94.0 (92.3,95.6)
	0.250	95.2 (91.2,98.4)	94.6 (93.1,96.2)
	0.300	93.6 (88.8,97.6)	95.6 (94.1,96.9)
	0.350	89.6 (84.0,94.4)	96.6 (95.3,97.8)
	0.400	87.2 (81.6,92.8)	97.5 (96.4,98.6)
Vertebral compression fracture	0.460	93.4 (90.1,96.0)	85.8 (82.1,89.6)
	0.500	92.6 (89.3,95.6)	90.9 (87.7,93.7)
	0.550	87.1 (83.1,90.8)	94.7 (91.8,96.9)

CTB: Sensitivity and specificity by operating point

Finding	Slice thickness range	Operating point	Sensitivity % (95% CI)	Specificity (95% CI)
Acute subarachnoid hemorrhage	≤1.5mm	0.014372	98.0 (95.2, 100.0)	89.4 (82.4, 95.3)
		0.060162	93.9 (89.8, 97.3)	96.5 (91.8, 100.0)
		0.082652	89.8 (85.0, 94.6)	100.0 (100.0, 100.0)
	>1.5mm & ≤5.0mm	0.020255	90.7 (86.3, 95.1)	92.4 (86.7, 97.1)
		0.030010	87.4 (82.4, 91.8)	96.2 (92.4, 99.0)
Acute subdural/epidural hematoma	≤1.5mm	0.060177	91.4 (88.1, 94.4)	86.7 (79.6, 92.9)
		0.101143	89.1 (85.5, 92.4)	94.9 (89.8, 99.0)
		0.135700	86.5 (82.5, 90.1)	96.9 (92.9, 100.0)
	>1.5mm & ≤5.0mm	0.060177	82.4 (78.6, 86.1)	89.6 (83.7, 94.8)
Intra-axial hemorrhage	≤1.5mm	0.322700	93.1 (90.8, 95.2)	85.6 (81.1, 89.6)
	>1.5mm & ≤5.0mm	0.203600	93.4 (91.3, 95.1)	85.1 (80.9, 88.9)
		0.322700	90.3 (87.9, 92.5)	90.3 (86.8, 93.8)
Intraventricular hemorrhage	≤1.5mm	0.015487	95.9 (90.4, 100.0)	90.9 (84.4, 97.4)
		0.051859	90.4 (83.6, 97.3)	97.4 (93.5, 100.0)
	>1.5mm & ≤5.0mm	0.008430	95.6 (91.2, 98.9)	86.0 (78.5, 92.5)
		0.015487	92.3 (86.8, 96.7)	89.2 (82.8, 94.6)
		0.051859	87.9 (80.2, 94.5)	97.8 (94.6, 100.0)
Mass effect	≤1.5mm	0.160195	97.0 (95.3, 98.4)	88.7 (83.5, 94.0)
		0.221484	96.6 (94.9, 98.2)	89.5 (84.2, 94.0)
	>1.5mm & ≤5.0mm	0.120944	96.8 (95.3, 98.1)	89.3 (84.5, 93.5)
		0.160195	95.3 (93.6, 97.0)	92.9 (88.7, 96.4)
Obstructive hydrocephalus	≤1.5mm	0.149943	97.3 (93.3, 100.0)	94.0 (89.0, 98.0)
		0.185900	94.7 (89.3, 98.7)	95.0 (90.0, 99.0)
		0.281473	92.0 (85.3, 97.3)	97.0 (93.0, 100.0)
	>1.5mm & ≤5.0mm	0.100591	97.6 (94.0, 100.0)	95.3 (90.7, 99.1)
		0.149943	95.2 (90.5, 98.8)	95.3 (90.7, 99.1)
		0.185900	94.0 (89.3, 98.8)	95.3 (90.7, 99.1)
		0.281473	88.1 (81.0, 94.0)	95.3 (90.7, 99.1)
Vasogenic edema	≤1.5mm	0.060584	91.7 (85.0, 98.3)	89.7 (83.2, 95.3)
		0.094076	90.0 (81.7, 96.7)	90.7 (85.0, 96.3)
		0.145352	90.0 (81.7, 96.7)	93.5 (88.8, 97.2)
	>1.5mm & ≤5.0mm	0.060584	94.9 (89.9, 99.0)	93.3 (89.7, 96.4)
		0.094076	93.9 (88.9, 98.0)	94.3 (90.7, 97.4)
		0.145352	90.9 (84.8, 96.0)	95.4 (92.3, 97.9)
		0.261255	89.9 (83.8, 94.9)	97.4 (94.8, 99.5)

Patient age

Results

The following tables outline patient age (younger to midlife adults aged ≤ 65 years versus elderly adults aged > 65 years) by sensitivity and specificity.

See:

- [CXR: Patient age – sensitivity and specificity](#), below
- [CTB: Patient age – sensitivity and specificity](#), page 12

CXR: Patient age – sensitivity and specificity

Finding	Operating point	Patient age	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion	0.380	≤ 65 years	98.0 (95.4,100.0)	88.3 (84.5,92.0)
		> 65 years	96.1 (94.0,98.2)	85.2 (80.6,89.7)
	0.425	≤ 65 years	94.7 (90.7,98.0)	90.2 (86.4,93.6)
		> 65 years	94.3 (91.5,96.7)	88.9 (84.8,92.6)
	0.450	≤ 65 years	92.1 (87.4,96.0)	92.0 (88.6,95.1)
		> 65 years	93.4 (90.3,95.8)	90.5 (86.8,94.2)
	0.475	≤ 65 years	89.4 (84.1,94.0)	94.3 (91.3,97.0)
		> 65 years	90.0 (86.7,93.1)	93.0 (89.7,95.9)
	0.500	≤ 65 years	88.7 (83.4,93.4)	96.2 (93.9,98.5)
		> 65 years	87.0 (83.1,90.3)	94.7 (91.8,97.1)
Pneumoperitoneum	0.250	≤ 65 years	100.0 (100.0,100.0)	88.4 (81.1,94.7)
		> 65 years	92.5 (84.9,98.1)	87.4 (81.1,93.7)
	0.300	≤ 65 years	98.1 (94.2,100.0)	92.6 (87.4,97.9)
		> 65 years	90.6 (83.0,98.1)	88.4 (82.1,94.7)
	0.350	≤ 65 years	96.2 (90.4,100.0)	94.7 (90.5,98.9)
		> 65 years	88.7 (79.2,96.2)	92.6 (87.4,97.9)
	0.400	≤ 65 years	96.2 (90.4,100.0)	95.8 (91.6,98.9)
		> 65 years	86.8 (77.4,94.3)	95.8 (91.6,98.9)
	0.450	≤ 65 years	94.2 (86.5,100.0)	97.9 (94.7,100.0)
		> 65 years	81.1 (69.8,90.6)	98.9 (96.8,100.0)

Finding	Operating point	Patient age	Sensitivity % (95% CI)	Specificity % (95% CI)
Pneumothorax	0.200	≤ 65 years	98.1 (96.3,99.5)	87.1 (83.2,90.7)
		> 65 years	96.1 (93.1,98.5)	89.3 (85.5,93.1)
	0.250	≤ 65 years	97.2 (94.9,99.1)	91.8 (88.6,95.0)
		> 65 years	95.1 (92.1,97.5)	92.0 (88.5,95.0)
	0.300	≤ 65 years	96.3 (93.5,98.6)	93.6 (90.7,96.4)
		> 65 years	93.6 (90.1,96.6)	94.7 (92.0,96.9)
	0.350	≤ 65 years	94.4 (91.2,97.2)	95.7 (93.2,97.9)
		> 65 years	91.6 (87.2,95.1)	95.4 (92.7,97.7)
	0.400	≤ 65 years	93.5 (90.2,96.7)	97.1 (95.4,98.9)
		> 65 years	87.7 (82.8,92.1)	96.2 (93.5,98.1)
Tension pneumothorax	0.225	≤ 65 years	96.1 (90.9,100.0)	93.1 (90.7,95.5)
		> 65 years	95.8 (89.6,100.0)	95.0 (92.6,96.9)
	0.250	≤ 65 years	96.1 (90.9,100.0)	93.8 (91.4,96.2)
		> 65 years	93.8 (87.5,100.0)	95.4 (93.3,97.4)
	0.300	≤ 65 years	94.8 (89.6,98.7)	95.5 (93.3,97.4)
		> 65 years	91.7 (83.3,97.9)	95.7 (93.5,97.6)
	0.350	≤ 65 years	89.6 (81.8,96.1)	96.9 (95.0,98.6)
		> 65 years	89.6 (79.2,95.9)	96.4 (94.5,98.1)
	0.400	≤ 65 years	88.3 (80.5,94.8)	98.3 (97.1,99.5)
		> 65 years	85.4 (75.0,93.8)	96.6 (94.7,98.3)
Vertebral compression fracture	0.460	≤ 65 years	92.3 (84.6,98.1)	91.5 (87.5,95.5)
		> 65 years	93.6 (90.5,96.4)	78.9 (71.8,85.2)
	0.500	≤ 65 years	92.3 (84.6,98.1)	94.3 (90.9,97.7)
		> 65 years	92.7 (89.1,95.9)	86.6 (81.0,91.6)
	0.550	≤ 65 years	80.8 (69.2,90.4)	96.6 (93.8,98.9)
		> 65 years	88.6 (84.5,92.3)	92.3 (87.3,96.5)

CTB: Patient age – sensitivity and specificity

Finding	Slice thickness range	Operating point	Patient age	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	≤1.5mm	0.014372	≤ 65 years	100.0 (100.0,100.0)	87.8 (79.6,95.9)
			> 65 years	97.2 (93.5,100.0)	91.7 (83.3,100.0)
		0.060162	≤ 65 years	95.0 (87.5,100.0)	93.9 (85.7,100.0)
			> 65 years	93.5 (88.8,97.2)	100.0 (100.0,100.0)
		0.082652	≤ 65 years	95.0 (87.5,100.0)	100.0 (100.0,100.0)
			> 65 years	87.9 (81.3,93.5)	100.0 (100.0,100.0)
	>1.5mm & ≤5.0mm	0.020255	≤ 65 years	94.3 (86.8,100.0)	91.1 (82.1,98.2)
			> 65 years	89.1 (83.7,94.6)	93.9 (85.7,100.0)
		0.030010	≤ 65 years	90.6 (81.1,96.2)	96.4 (91.1,100.0)
			> 65 years	86.0 (79.8,91.5)	95.9 (89.8,100.0)
Acute subdural/epidural hematoma	≤1.5mm	0.060177	≤ 65 years	90.7 (85.2,95.4)	92.7 (82.9,100.0)
			> 65 years	91.8 (87.7,95.4)	82.5 (71.9,91.2)
		0.101143	≤ 65 years	86.1 (79.6,92.6)	100.0 (100.0,100.0)
			> 65 years	90.8 (86.7,94.9)	91.2 (82.5,98.2)
		0.135700	≤ 65 years	80.6 (73.1,87.0)	100.0 (100.0,100.0)
			> 65 years	89.7 (85.1,93.8)	94.7 (87.7,100.0)
	>1.5mm & ≤5.0mm	0.060177	≤ 65 years	77.1 (69.4,84.0)	93.1 (86.2,98.3)
			> 65 years	85.4 (81.0,89.3)	87.0 (77.9,93.5)
Intra-axial hemorrhage	≤1.5mm	0.322700	≤ 65 years	91.7 (87.5,95.3)	88.8 (82.7,94.9)
			> 65 years	94.1 (91.3,96.5)	83.1 (76.6,89.5)
	>1.5mm & ≤5.0mm	0.203600	≤ 65 years	93.8 (90.2,96.9)	85.2 (78.7,91.8)
			> 65 years	93.1 (90.4,95.6)	84.9 (79.5,90.4)
		0.322700	≤ 65 years	90.6 (86.6,94.2)	90.2 (84.4,95.1)
			> 65 years	90.1 (86.8,93.1)	90.4 (85.5,94.6)
Intraventricular hemorrhage	≤1.5mm	0.015487	≤ 65 years	95.8 (87.5,100.0)	100.0 (100.0,100.0)
			> 65 years	95.9 (89.8,100.0)	85.1 (74.5,93.6)
		0.051859	≤ 65 years	87.5 (75.0,100.0)	100.0 (100.0,100.0)
			> 65 years	91.8 (83.7,98.0)	95.7 (89.4,100.0)
	>1.5mm & ≤5.0mm	0.008430	≤ 65 years	96.7 (90.0,100.0)	89.5 (78.9,97.4)
			> 65 years	95.1 (88.5,100.0)	83.6 (72.7,92.7)
		0.015487	≤ 65 years	96.7 (90.0,100.0)	94.7 (86.8,100.0)
			> 65 years	90.2 (82.0,96.7)	85.5 (76.4,94.5)
		0.051859	≤ 65 years	86.7 (73.3,96.7)	100.0 (100.0,100.0)
			> 65 years	88.5 (80.3,96.7)	96.4 (90.9,100.0)

Finding	Slice thickness range	Operating point	Patient age	Sensitivity % (95% CI)	Specificity % (95% CI)
Mass effect	≤1.5mm	0.160195	≤ 65 years	97.5 (94.9,99.5)	91.7 (83.3,97.9)
			> 65 years	96.6 (94.6,98.6)	87.1 (80.0,94.1)
		0.221484	≤ 65 years	97.5 (94.9,99.5)	91.7 (83.3,97.9)
			> 65 years	95.9 (93.6,98.0)	88.2 (81.2,94.1)
	>1.5mm & ≤5.0mm	0.120944	≤ 65 years	97.1 (95.1,98.8)	94.1 (88.2,98.5)
			> 65 years	96.6 (94.6,98.3)	86.0 (80.0,92.0)
		0.160195	≤ 65 years	95.5 (92.6,98.0)	97.1 (92.6,100.0)
			> 65 years	95.1 (92.9,97.1)	90.0 (84.0,95.0)
Obstructive hydrocephalus	≤1.5mm	0.149943	≤ 65 years	95.3 (88.4,100.0)	91.8 (83.7,98.0)
			> 65 years	100.0 (100.0,100.0)	96.1 (90.2,100.0)
		0.185900	≤ 65 years	93.0 (83.7,100.0)	91.8 (83.7,98.0)
			> 65 years	96.9 (90.6,100.0)	98.0 (94.1,100.0)
		0.281473	≤ 65 years	93.0 (83.7,100.0)	95.9 (89.8,100.0)
			> 65 years	90.6 (78.1,100.0)	98.0 (94.1,100.0)
	>1.5mm & ≤5.0mm	0.100591	≤ 65 years	97.8 (93.5,100.0)	92.3 (84.6,98.1)
			> 65 years	97.4 (92.1,100.0)	98.2 (94.5,100.0)
		0.149943	≤ 65 years	95.7 (89.1,100.0)	92.3 (84.6,98.1)
			> 65 years	94.7 (86.8,100.0)	98.2 (94.5,100.0)
		0.185900	≤ 65 years	93.5 (84.8,100.0)	92.3 (84.6,98.1)
			> 65 years	94.7 (86.8,100.0)	98.2 (94.5,100.0)
		0.281473	≤ 65 years	91.3 (82.6,97.8)	92.3 (84.6,98.1)
			> 65 years	84.2 (73.7,94.7)	98.2 (94.5,100.0)
Vasogenic edema	≤1.5mm	0.060584	≤65 years	92.0 (80.0,100.0)	84.3 (74.5,94.1)
			> 65 years	91.4 (80.0,100.0)	94.6 (87.5,100.0)
		0.094076	≤ 65 years	88.0 (76.0,100.0)	86.3 (76.5,94.1)
			> 65 years	91.4 (80.0,100.0)	94.6 (87.5,100.0)
		0.145352	≤ 65 years	88.0 (76.0,100.0)	90.2 (82.4,98.0)
			> 65 years	91.4 (80.0,100.0)	96.4 (91.1,100.0)
	>1.5mm & ≤5.0mm	0.060584	≤ 65 years	95.0 (87.5,100.0)	90.8 (83.9,96.6)
			> 65 years	94.9 (88.1,100.0)	95.3 (91.6,99.1)
		0.094076	≤ 65 years	92.5 (82.5,100.0)	93.1 (87.4,97.7)
			> 65 years	94.9 (88.1,100.0)	95.3 (91.6,99.1)
		0.145352	≤ 65 years	90.0 (80.0,97.5)	94.3 (89.7,98.9)
			> 65 years	91.5 (83.1,98.3)	96.3 (92.5,99.1)
		0.261255	≤ 65 years	87.5 (77.5,97.5)	96.6 (92.0,100.0)
			> 65 years	91.5 (83.1,98.3)	98.1 (95.3,100.0)

Patient sex

Results

The following tables outline the sensitivity and specificity for each finding by operating point and patient sex.

See:

- [CXR: Patient sex - sensitivity and specificity](#), below
- [CTB: Patient sex - sensitivity and specificity](#), page 16

CXR: Patient sex - sensitivity and specificity

Finding	Operating point	Sex	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion	0.380	Female	97.8 (95.5,99.6)	86.6 (82.3,90.6)
		Male	95.7 (93.0,98.1)	87.0 (83.0,90.9)
	0.425	Female	94.6 (91.5,97.3)	89.4 (85.0,92.9)
		Male	94.2 (91.1,96.9)	89.7 (86.2,93.3)
	0.450	Female	92.0 (88.4,95.5)	90.6 (86.6,94.1)
		Male	93.8 (90.7,96.5)	92.1 (88.9,95.3)
	0.475	Female	88.4 (84.4,92.4)	92.5 (89.0,95.3)
		Male	91.1 (87.6,94.6)	94.9 (92.1,97.6)
	0.500	Female	87.1 (82.6,91.5)	94.9 (92.1,97.2)
		Male	88.0 (84.1,91.9)	96.0 (93.7,98.4)
Pneumoperitoneum	0.250	Female	98.0 (93.9,100.0)	87.2 (79.8,93.6)
		Male	94.6 (87.5,100.0)	88.5 (82.3,94.8)
	0.300	Female	95.9 (89.8,100.0)	90.4 (84.0,95.7)
		Male	92.9 (85.7,98.2)	90.6 (84.4,95.8)
	0.350	Female	93.9 (85.7,100.0)	92.6 (87.2,97.9)
		Male	91.1 (83.9,98.2)	94.8 (89.6,99.0)
	0.400	Female	91.8 (83.7,98.0)	94.7 (89.4,98.9)
		Male	91.1 (83.9,98.2)	96.9 (92.7,100.0)
	0.450	Female	85.7 (75.5,93.9)	97.9 (94.7,100.0)
		Male	89.3 (80.4,96.4)	99.0 (96.9,100.0)

Finding	Operating point	Sex	Sensitivity % (95% CI)	Specificity % (95% CI)
Pneumothorax	0.200	Female	96.6 (93.9,98.9)	89.2 (85.2,93.2)
		Male	97.5 (95.4,99.2)	87.3 (83.2,91.1)
	0.250	Female	96.1 (92.7,98.3)	92.8 (89.6,96.0)
		Male	96.2 (93.7,98.3)	91.1 (87.7,94.2)
	0.300	Female	95.0 (91.6,97.8)	95.6 (92.8,98.0)
		Male	95.0 (92.1,97.5)	92.8 (89.7,95.5)
	0.350	Female	92.2 (88.3,96.1)	96.8 (94.4,98.8)
		Male	93.7 (90.8,96.7)	94.5 (92.1,96.9)
	0.400	Female	89.4 (84.4,93.9)	98.4 (96.8,99.6)
		Male	91.6 (87.9,95.0)	95.2 (92.8,97.6)
Tension Pneumothorax	0.225	Female	94.5 (87.3,100.0)	94.7 (92.5,96.8)
		Male	97.1 (92.9,100.0)	93.5 (91.1,95.7)
	0.250	Female	94.5 (87.3,100.0)	95.5 (93.3,97.3)
		Male	95.7 (91.4,100.0)	93.9 (91.8,96.1)
	0.300	Female	94.5 (87.3,100.0)	96.0 (94.1,97.9)
		Male	92.9 (87.1,98.6)	95.2 (93.3,97.2)
	0.350	Female	90.9 (83.6,98.2)	97.3 (95.5,98.9)
		Male	88.6 (81.4,95.7)	96.1 (94.4,97.8)
	0.400	Female	90.9 (83.6,98.2)	97.6 (96.0,98.9)
		Male	84.3 (75.7,91.4)	97.4 (95.9,98.7)
Vertebral compression fracture	0.460	Female	94.4 (91.0,97.2)	85.3 (79.4,90.6)
		Male	91.5 (85.1,96.8)	86.5 (81.1,91.2)
	0.500	Female	93.8 (90.4,97.2)	92.4 (88.2,95.9)
		Male	90.4 (84.0,95.7)	89.2 (84.5,93.9)
	0.550	Female	88.2 (83.1,92.7)	94.7 (91.2,97.6)
		Male	85.1 (77.7,91.5)	94.6 (90.5,98.0)

CTB: Patient sex – sensitivity and specificity

Finding	Slice thickness range	Operating point	Sex	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	≤1.5mm	0.014372	Female	97.7 (94.2,100.0)	88.6 (77.1,97.1)
			Male	98.4 (95.1,100.0)	90.0 (82.0,98.0)
		0.060162	Female	93.0 (87.2,97.7)	97.1 (91.4,100.0)
			Male	95.1 (88.5,100.0)	96.0 (90.0,100.0)
		0.082652	Female	88.4 (80.2,94.2)	100.0 (100.0,100.0)
			Male	91.8 (83.6,98.4)	100.0 (100.0,100.0)
	>1.5mm & ≤5.0mm	0.020255	Female	88.5 (81.2,94.8)	93.9 (85.7,100.0)
			Male	93.0 (87.2,97.7)	91.1 (82.1,98.2)
		0.030010	Female	84.4 (76.0,90.6)	95.9 (89.8,100.0)
			Male	90.7 (83.7,96.5)	96.4 (91.1,100.0)
Acute subdural/epidural hematoma	≤1.5mm	0.060177	Female	88.5 (83.2,93.9)	86.5 (76.9,94.2)
			Male	93.6 (89.5,97.1)	87.0 (76.1,95.7)
		0.101143	Female	86.3 (80.2,91.6)	94.2 (88.5,100.0)
			Male	91.3 (86.6,95.3)	95.7 (89.1,100.0)
		0.135700	Female	84.7 (77.9,90.8)	96.2 (90.4,100.0)
			Male	87.8 (83.1,92.4)	97.8 (93.5,100.0)
	>1.5mm & ≤5.0mm	0.060177	Female	80.2 (73.8,86.0)	91.0 (83.6,97.0)
			Male	84.0 (79.1,88.4)	88.2 (80.9,95.6)
Intra-axial hemorrhage	≤1.5mm	0.322700	Female	92.5 (88.8,95.8)	88.7 (82.6,93.9)
			Male	93.6 (90.6,96.2)	82.2 (74.8,88.8)
	>1.5mm & ≤5.0mm	0.203600	Female	93.0 (89.9,96.1)	86.2 (80.3,91.4)
			Male	93.6 (90.9,96.0)	83.8 (77.2,89.7)
		0.322700	Female	90.3 (86.4,93.8)	89.5 (84.2,94.1)
			Male	90.3 (86.9,93.3)	91.2 (86.0,95.6)
Intraventricular hemorrhage	≤1.5mm	0.015487	Female	100.0 (100.0,100.0)	94.4 (86.1,100.0)
			Male	90.9 (78.8,100.0)	87.8 (78.0,97.6)
		0.051859	Female	97.5 (92.5,100.0)	100.0 (100.0,100.0)
			Male	81.8 (66.7,93.9)	95.1 (87.8,100.0)
	>1.5mm & ≤5.0mm	0.008430	Female	100.0 (100.0,100.0)	88.6 (79.5,97.7)
			Male	91.1 (82.2,97.8)	83.7 (73.5,91.8)
		0.015487	Female	95.7 (89.1,100.0)	90.9 (81.8,97.7)
			Male	88.9 (80.0,97.8)	87.8 (77.6,95.9)
		0.051859	Female	91.3 (82.6,97.8)	100.0 (100.0,100.0)
			Male	84.4 (73.3,93.3)	95.9 (89.8,100.0)

Finding	Slice thickness range	Operating point	Sex	Sensitivity % (95% CI)	Specificity % (95% CI)
Mass effect	≤1.5mm	0.160195	Female	98.3 (96.6,99.6)	90.8 (83.1,96.9)
			Male	95.7 (93.0,98.0)	86.8 (77.9,94.1)
		0.221484	Female	97.5 (95.4,99.2)	90.8 (83.1,96.9)
			Male	95.7 (93.0,98.0)	88.2 (79.4,95.6)
	>1.5mm & ≤5.0mm	0.120944	Female	96.7 (94.5,98.5)	85.9 (78.2,93.6)
			Male	96.9 (94.7,98.8)	92.2 (86.7,96.7)
		0.160195	Female	94.5 (91.5,97.1)	89.7 (82.1,96.2)
			Male	96.0 (93.8,98.1)	95.6 (91.1,98.9)
Obstructive hydrocephalus	≤1.5mm	0.149943	Female	95.5 (88.6,100.0)	92.6 (85.2,98.1)
			Male	100.0 (100.0,100.0)	95.7 (89.1,100.0)
		0.185900	Female	90.9 (81.8,97.7)	94.4 (88.9,100.0)
			Male	100.0 (100.0,100.0)	95.7 (89.1,100.0)
		0.281473	Female	86.4 (75.0,95.5)	96.3 (90.7,100.0)
			Male	100.0 (100.0,100.0)	97.8 (93.5,100.0)
	>1.5mm & ≤5.0mm	0.100591	Female	96.0 (90.0,100.0)	94.5 (87.3,100.0)
			Male	100.0 (100.0,100.0)	96.2 (90.4,100.0)
		0.149943	Female	92.0 (84.0,98.0)	94.5 (87.3,100.0)
			Male	100.0 (100.0,100.0)	96.2 (90.4,100.0)
		0.185900	Female	90.0 (82.0,98.0)	94.5 (87.3,100.0)
			Male	100.0 (100.0,100.0)	96.2 (90.4,100.0)
		0.281473	Female	84.0 (74.0,94.0)	94.5 (87.3,100.0)
			Male	94.1 (85.3,100.0)	96.2 (90.4,100.0)
Vasogenic edema	≤1.5mm	0.060584	Female	92.9 (82.1,100.0)	89.8 (81.6,98.0)
			Male	90.6 (78.1,100.0)	89.7 (81.0,96.6)
		0.094076	Female	92.9 (82.1,100.0)	89.8 (81.6,98.0)
			Male	87.5 (75.0,96.9)	91.4 (84.5,98.3)
		0.145352	Female	92.9 (82.1,100.0)	93.9 (85.7,100.0)
			Male	87.5 (75.0,96.9)	93.1 (86.2,98.3)
	>1.5mm & ≤5.0mm	0.060584	Female	97.8 (93.3,100.0)	94.0 (89.0,98.0)
			Male	92.6 (85.2,98.1)	92.6 (87.2,97.9)
		0.094076	Female	97.8 (93.3,100.0)	95.0 (90.0,99.0)
			Male	90.7 (83.3,98.1)	93.6 (88.3,97.9)
		0.145352	Female	95.6 (88.9,100.0)	95.0 (90.0,99.0)
			Male	87.0 (77.8,96.3)	95.7 (91.5,98.9)
		0.261255	Female	95.6 (88.9,100.0)	97.0 (93.0,100.0)
			Male	85.2 (75.9,94.4)	97.9 (94.7,100.0)

Patient race and ethnicity

Results

The following tables outline the sensitivity and specificity for each finding by patient race subgroup and the distribution for each finding by patient ethnicity subgroup.

The analysis used a single representative operating point, with data organized into the following race subgroups:

- White or Caucasian
- All other groups (includes American Indian or Alaskan Native, Asian, Black or African American, Native Hawaiian or Other Pacific Islander, Other, Two or more races)*, and
- Unknown (includes Declined or Unavailable).

This analysis suggests that the device performance is not distinctly different between race subgroups.

* These groups are listed separately for CXR.

See:

- [CXR: Patient race – sensitivity and specificity, page 19](#)
- [CXR: Patient race - distribution, page 20](#)
- [CXR: Patient ethnicity - distribution, page 21](#)
- [CTB: Patient race – sensitivity and specificity, page 22](#)
- [CTB: Patient race - distribution, page 24](#)
- [CTB: Patient ethnicity - distribution, page 31](#)

CXR: Patient race – sensitivity and specificity

Finding	Operating point	Race	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion	0.450	Declined or Unavailable	75.0 (37.5,100.0)	81.2 (62.5,100.0)
		American Indian or Alaska Native	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Asian	94.4 (83.3,100.0)	100.0 (100.0,100.0)
		Black or African American	77.4 (61.3,90.3)	92.7 (82.9,100.0)
		White or Caucasian	94.1 (91.6,96.3)	91.1 (88.2,93.8)
		Other	100.0 (100.0,100.0)	90.0 (75.0,100.0)
		Two or more	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Pneumoperitoneum	0.300	Declined or Unavailable	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		American Indian or Alaska Native	-	100.0 (100.0,100.0)
		Asian	100.0 (100.0,100.0)	85.7 (57.1,100.0)
		Black or African American	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		White or Caucasian	92.9 (87.1,97.6)	88.8 (83.6,93.4)
		Other	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Two or more	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Pneumothorax	0.300	Declined or Unavailable	94.1 (82.4,100.0)	100.0 (100.0,100.0)
		American Indian or Alaska Native	-	-
		Asian	100.0 (100.0,100.0)	88.2 (70.6,100.0)
		Black or African American	88.2 (70.6,100.0)	100.0 (100.0,100.0)
		White or Caucasian	94.8 (92.5,97.1)	93.2 (90.7,95.5)
		Other	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Two or more	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Tension pneumothorax	0.300	Declined or Unavailable	88.9 (66.7,100.0)	89.3 (78.6,100.0)
		American Indian or Alaska Native	-	-
		Asian	100.0 (100.0,100.0)	85.2 (70.4,96.3)
		Black or African American	100.0 (100.0,100.0)	97.6 (92.9,100.0)
		White or Caucasian	92.9 (87.8,98.0)	96.0 (94.5,97.4)
		Other	100.0 (100.0,100.0)	97.0 (90.9,100.0)
		Two or more	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Vertebral compression fracture	0.460	Declined or Unavailable	75.0 (25.0,100.0)	88.9 (66.7,100.0)
		American Indian or Alaska Native	100.0 (100.0,100.0)	-
		Asian	100.0 (100.0,100.0)	87.5 (62.5,100.0)
		Black or African American	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		White or Caucasian	94.3 (91.5,97.2)	84.3 (79.6,88.6)
		Other	75.0 (50.0,100.0)	85.7 (64.3,100.0)
		Two or more	-	83.3 (44.2,100.0)

CXR: Patient race - distribution

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Pleural effusion		(N=482	(N=507	(N=989
	American Indian or Alaska Native	1 (0.2%)	2 (0.4%)	3 (0.3%)
	Asian	18 (3.7%)	17 (3.4%)	35 (3.5%)
	Black or African American	31 (6.4%)	41 (8.1%)	72 (7.3%)
	White or Caucasian	406 (84.2%)	406 (80.1%)	812 (82.1%)
	Other	16 (3.3%)	20 (3.9%)	36 (3.6%)
	Two or more	2 (0.4%)	5 (1.0%)	7 (0.7%)
	Declined or Unavailable	8 (1.7%)	16 (3.2%)	24 (2.4%)
Pneumoperitoneum		(N=105	(N=190	(N=295
	American Indian or Alaska Native	0 (0.0%)	1 (0.5%)	1 (0.3%)
	Asian	5 (4.8%)	7 (3.7%)	12 (4.1%)
	Black or African American	6 (5.7%)	11 (5.8%)	17 (5.8%)
	White or Caucasian	85 (81.0%)	152 (80.0%)	237 (80.3%)
	Other	5 (4.8%)	12 (6.3%)	17 (5.8%)
	Two or more	1 (1.0%)	3 (1.6%)	4 (1.4%)
	Declined or Unavailable	3 (2.9%)	4 (2.1%)	7 (2.4%)
Pneumothorax		(N = 418)	(N = 542)	(N = 960)
	American Indian or Alaska Native	0 (0.0%)	0 (0.0%)	0 (0.0%)
	Asian	13 (3.1%)	17 (3.1%)	30 (3.1%)
	Black or African American	17 (4.1%)	31 (5.7%)	48 (5.0%)
	White or Caucasian	348 (83.3%)	443 (81.7%)	791 (82.4%)
	Other	22 (5.3%)	19 (3.5%)	41 (4.3%)
	Two or more	1 (0.2%)	12 (2.2%)	13 (1.4%)
	Declined or Unavailable	17 (4.1%)	20 (3.7%)	37 (3.9%)
Tension pneumothorax		(N=125	(N=835	(N=960
	American Indian or Alaska Native	0 (0.0%)	0 (0.0%)	0 (0.0%)
	Asian	3 (2.4%)	27 (3.2%)	30 (3.1%)
	Black or African American	6 (4.8%)	42 (5.0%)	48 (5.0%)
	White or Caucasian	98 (78.4%)	693 (83.0%)	791 (82.4%)
	Other	8 (6.4%)	33 (4.0%)	41 (4.3%)
	Two or more	1 (0.8%)	12 (1.4%)	13 (1.4%)
	Declined or Unavailable	9 (7.2%)	28 (3.4%)	37 (3.9%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Vertebral compression fracture		(N=272	(N=318	(N=590
	American Indian or Alaska Native	1 (0.4%)	0 (0.0%)	1 (0.2%)
	Asian	5 (1.8%)	8 (2.5%)	13 (2.2%)
	Black or African American	4 (1.5%)	26 (8.2%)	30 (5.1%)
	White or Caucasian	246 (90.4%)	255 (80.2%)	501 (84.9%)
	Other	12 (4.4%)	14 (4.4%)	26 (4.4%)
	Two or more	0 (0.0%)	6 (1.9%)	6 (1.0%)
	Declined or Unavailable	4 (1.5%)	9 (2.8%)	13 (2.2%)

CXR: Patient ethnicity - distribution

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Pleural effusion		(N=482	(N=507	(N=989
	Hispanic	24 (5.0%)	40 (7.9%)	64 (6.5%)
	Not Hispanic	444 (92.1%)	455 (89.7%)	899 (90.9%)
	Declined or Unavailable	14 (2.9%)	12 (2.4%)	26 (2.6%)
Pneumoperitoneum		(N=105	(N=190	(N=295
	Hispanic	5 (4.8%)	19 (10.0%)	24 (8.1%)
	Not Hispanic	95 (90.5%)	158 (83.2%)	253 (85.8%)
	Declined or Unavailable	5 (4.8%)	13 (6.8%)	18 (6.1%)
Pneumothorax		(N = 418)	(N = 542)	(N = 960)
	Hispanic	25 (6.0%)	45 (8.3%)	70 (7.3%)
	Not Hispanic	369 (88.3%)	459 (84.7%)	828 (86.2%)
	Declined or Unavailable	24 (5.7%)	38 (7.0%)	62 (6.5%)
Tension pneumothorax		(N=125	(N=835	(N=960
	Hispanic	12 (9.6%)	58 (6.9%)	70 (7.3%)
	Not Hispanic	100 (80.0%)	728 (87.2%)	828 (86.2%)
	Declined or Unavailable	13 (10.4%)	49 (5.9%)	62 (6.5%)
Vertebral compression fracture		(N=272	(N=318	(N=590
	Hispanic	17 (6.2%)	32 (10.1%)	49 (8.3%)
	Not Hispanic	246 (90.4%)	270 (84.9%)	516 (87.5%)
	Declined or Unavailable	9 (3.3%)	16 (5.0%)	25 (4.2%)

CTB: Patient race – sensitivity and specificity

Finding	Operating point	Race	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	0.060162	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	93.3 (88.3,97.5)	96.8 (92.1,100.0)
		All other groups	100.0 (100.0,100.0)	93.8 (81.2,100.0)
		Unknown	90.0 (70.0,100.0)	100.0 (100.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	81.9 (75.8,87.2)	97.5 (93.8,100.0)
		All other groups	90.9 (77.3,100.0)	100.0 (100.0,100.0)
		Unknown	90.9 (72.7,100.0)	100.0 (100.0,100.0)
Acute subdural/epidural hematoma	0.060177	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	90.8 (87.0,94.5)	85.7 (77.1,92.9)
		All other groups	92.2 (84.3,98.0)	87.5 (70.8,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	81.9 (77.9,86.0)	88.2 (81.4,94.1)
		All other groups	81.7 (71.7,90.0)	93.1 (82.8,100.0)
		Unknown	93.8 (81.2,100.0)	100.0 (100.0,100.0)
Intra-axial hemorrhage	0.322700	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	93.1 (90.6,95.6)	85.2 (80.1,90.3)
		All other groups	92.9 (87.8,98.0)	88.9 (77.8,97.2)
		Unknown	95.0 (85.0,100.0)	80.0 (50.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	89.8 (87.2,92.5)	91.1 (86.8,94.5)
		All other groups	91.2 (85.0,96.5)	92.7 (85.4,100.0)
		Unknown	95.5 (86.4,100.0)	66.7 (41.7,91.7)
Intraventricular hemorrhage	0.015487	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	98.2 (94.7,100.0)	93.3 (86.7,98.3)
		All other groups	85.7 (64.3,100.0)	80.0 (60.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	93.1 (86.1,98.6)	89.9 (83.5,96.2)
		All other groups	86.7 (66.7,100.0)	83.3 (58.3,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Race	Sensitivity % (95% CI)	Specificity % (95% CI)
Mass effect	0.160195	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	96.7 (94.9,98.5)	87.5 (80.8,93.3)
		All other groups	97.5 (93.8,100.0)	95.8 (87.5,100.0)
		Unknown	100.0 (100.0,100.0)	80.0 (40.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	94.8 (92.8,96.7)	92.1 (87.1,95.7)
		All other groups	97.8 (94.6,100.0)	100.0 (100.0,100.0)
		Unknown	94.4 (83.3,100.0)	80.0 (40.0,100.0)
Obstructive hydrocephalus	0.149943	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	96.5 (91.2,100.0)	94.8 (89.6,98.7)
		All other groups	100.0 (100.0,100.0)	94.1 (82.4,100.0)
		Unknown	100.0 (100.0,100.0)	83.3 (50.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	93.8 (87.7,98.5)	96.3 (91.5,100.0)
		All other groups	100.0 (100.0,100.0)	94.7 (84.2,100.0)
		Unknown	100.0 (100.0,100.0)	83.3 (50.0,100.0)
Vasogenic edema	0.094076	Slice thickness range: $\leq 1.5\text{mm}$		
		White or Caucasian	93.5 (84.8,100.0)	89.0 (81.7,95.1)
		All other groups	75.0 (50.0,100.0)	94.7 (84.2,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$		
		White or Caucasian	96.2 (92.4,100.0)	93.5 (89.0,97.4)
		All other groups	82.4 (58.8,100.0)	96.9 (90.6,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)

CTB: Patient race - distribution

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Acute subarachnoid hemorrhage	≤1.5mm		N=232	N=18
		American Indian or Alaska Native	1 (0.4%)	0 (0.0%)
		Asian	12 (5.2%)	0 (0.0%)
		Black or African American	6 (2.6%)	1 (5.6%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	183 (78.9%)	5 (27.8%)
		All other groups	12 (5.2%)	7 (38.9%)
		Two or more races	2 (0.9%)	1 (5.6%)
		Declined	2 (0.9%)	0 (0.0%)
		Unavailable	14 (6.0%)	4 (22.2%)
	>1.5mm & ≤5.0mm		N=287	N=17
		American Indian or Alaska Native	2 (0.7%)	0 (0.0%)
		Asian	12 (4.2%)	0 (0.0%)
		Black or African American	9 (3.1%)	1 (5.9%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	230 (80.1%)	4 (23.5%)
		All other groups	12 (4.2%)	7 (41.2%)
		Two or more races	4 (1.4%)	1 (5.9%)
		Declined	3 (1.0%)	0 (0.0%)
		Unavailable	15 (5.2%)	4 (23.5%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Acute subdural/ epidural hematoma	≤1.5mm		N=401	N=42
		American Indian or Alaska	0 (0.0%)	0 (0.0%)
		Asian	20 (5.0%)	0 (0.0%)
		Black or African American	25 (6.2%)	2 (4.8%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	308 (76.8%)	14 (33.3%)
		All other groups	26 (6.5%)	20 (47.6%)
		Two or more races	4 (1.0%)	2 (4.8%)
		Declined	2 (0.5%)	1 (2.4%)
		Unavailable	16 (4.0%)	3 (7.1%)
	>1.5mm & ≤5.0mm		N=532	N=51
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	21 (3.9%)	0 (0.0%)
		Black or African American	27 (5.1%)	2 (3.9%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	423 (79.5%)	16 (31.4%)
		All other groups	33 (6.2%)	26 (51.0%)
		Two or more races	8 (1.5%)	3 (5.9%)
		Declined	3 (0.6%)	1 (2.0%)
		Unavailable	17 (3.2%)	3 (5.9%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Intra-axial hemorrhage	≤1.5mm		N=702	N=57
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	46 (6.6%)	0 (0.0%)
		Black or African American	42 (6.0%)	4 (7.0%)
		Native Hawaiian or Other Pacific Islander	1 (0.1%)	1 (1.8%)
		White or Caucasian	538 (76.6%)	13 (22.8%)
		All other groups	39 (5.6%)	33 (57.9%)
		Two or more races	6 (0.9%)	2 (3.5%)
		Declined	4 (0.6%)	0 (0.0%)
		Unavailable	26 (3.7%)	4 (7.0%)
	>1.5mm & ≤5.0mm		N=875	N=68
		American Indian or Alaska Native	1 (0.1%)	0 (0.0%)
		Asian	50 (5.7%)	0 (0.0%)
		Black or African American	51 (5.8%)	5 (7.4%)
		Native Hawaiian or Other Pacific Islander	2 (0.2%)	1 (1.5%)
		White or Caucasian	687 (78.5%)	19 (27.9%)
		All other groups	42 (4.8%)	36 (52.9%)
		Two or more races	8 (0.9%)	3 (4.4%)
		Declined	4 (0.5%)	0 (0.0%)
		Unavailable	30 (3.4%)	4 (5.9%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Intraventricular hemorrhage	≤1.5mm		N=150	N=17
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	4 (2.7%)	0 (0.0%)
		Black or African American	16 (10.7%)	3 (17.6%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	117 (78.0%)	5 (29.4%)
		All other groups	8 (5.3%)	7 (41.2%)
		Two or more races	1 (0.7%)	0 (0.0%)
		Declined	1 (0.7%)	0 (0.0%)
		Unavailable	3 (2.0%)	2 (11.8%)
	>1.5mm & ≤5.0mm		N=184	N=16
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	4 (2.2%)	0 (0.0%)
		Black or African American	16 (8.7%)	3 (18.8%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	151 (82.1%)	5 (31.2%)
		All other groups	7 (3.8%)	6 (37.5%)
		Two or more races	0 (0.0%)	0 (0.0%)
		Declined	1 (0.5%)	0 (0.0%)
		Unavailable	5 (2.7%)	2 (12.5%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Mass effect	≤1.5mm		N=626	N=50
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	22 (3.5%)	0 (0.0%)
		Black or African American	46 (7.3%)	4 (8.0%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	498 (79.6%)	19 (38.0%)
		All other groups	32 (5.1%)	24 (48.0%)
		Two or more races	5 (0.8%)	2 (4.0%)
		Declined	3 (0.5%)	0 (0.0%)
		Unavailable	20 (3.2%)	1 (2.0%)
	>1.5mm & ≤5.0mm		N=762	N=51
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	29 (3.8%)	0 (0.0%)
		Black or African American	46 (6.0%)	4 (7.8%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	623 (81.8%)	19 (37.3%)
		All other groups	34 (4.5%)	25 (49.0%)
		Two or more races	7 (0.9%)	2 (3.9%)
		Declined	3 (0.4%)	0 (0.0%)
		Unavailable	20 (2.6%)	1 (2.0%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Obstructive hydrocephalus	≤1.5mm		N=175	N=17
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	6 (3.4%)	0 (0.0%)
		Black or African American	12 (6.9%)	0 (0.0%)
		Native Hawaiian or Other Pacific Islander	1 (0.6%)	0 (0.0%)
		White or Caucasian	134 (76.6%)	5 (29.4%)
		All other groups	11 (6.3%)	9 (52.9%)
		Two or more races	2 (1.1%)	2 (11.8%)
		Declined	2 (1.1%)	0 (0.0%)
		Unavailable	7 (4.0%)	1 (5.9%)
	>1.5mm & ≤5.0mm		N=191	N=17
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	7 (3.7%)	0 (0.0%)
		Black or African American	14 (7.3%)	0 (0.0%)
		Native Hawaiian or Other Pacific Islander	1 (0.5%)	0 (0.0%)
		White or Caucasian	147 (77.0%)	5 (29.4%)
		All other groups	11 (5.8%)	9 (52.9%)
		Two or more races	2 (1.0%)	2 (11.8%)
		Declined	2 (1.0%)	0 (0.0%)
		Unavailable	7 (3.7%)	1 (5.9%)

Finding	Slice thickness range	Race	Total	Hispanic ethnicity
Vasogenic edema	≤1.5mm		N=167	N=10
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	11 (6.6%)	0 (0.0%)
		Black or African American	10 (6.0%)	0 (0.0%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	128 (76.6%)	2 (20.0%)
		All other groups	9 (5.4%)	7 (70.0%)
		Two or more races	1 (0.6%)	0 (0.0%)
		Declined	1 (0.6%)	0 (0.0%)
		Unavailable	7 (4.2%)	1 (10.0%)
	>1.5mm & ≤5.0mm		N=293	N=20
		American Indian or Alaska Native	0 (0.0%)	0 (0.0%)
		Asian	14 (4.8%)	0 (0.0%)
		Black or African American	18 (6.1%)	1 (5.0%)
		Native Hawaiian or Other Pacific Islander	0 (0.0%)	0 (0.0%)
		White or Caucasian	234 (79.9%)	6 (30.0%)
		All other groups	13 (4.4%)	10 (50.0%)
		Two or more races	4 (1.4%)	1 (5.0%)
		Declined	1 (0.3%)	0 (0.0%)
		Unavailable	9 (3.1%)	2 (10.0%)

CTB: Patient ethnicity - distribution

Finding	Slice thickness range	Race	Total
Acute subarachnoid hemorrhage	≤1.5mm		N=232
		Hispanic	18 (7.8%)
		Not Hispanic	195 (84.1%)
		Prefer not to say/Decline	2 (0.9%)
		Unavailable	17 (7.3%)
	>1.5mm & ≤5.0mm		N=287
		Hispanic	17 (5.9%)
		Not Hispanic	249 (86.8%)
		Prefer not to say/Decline	2 (0.7%)
		Unavailable	19 (6.6%)
Acute subdural/epidural hematoma	≤1.5mm		N=401
		Hispanic	42 (10.5%)
		Not Hispanic	337 (84.0%)
		Prefer not to say/Decline	0 (0.0%)
		Unavailable	22 (5.5%)
	>1.5mm & ≤5.0mm		N=532
		Hispanic	51 (9.6%)
		Not Hispanic	458 (86.1%)
		Prefer not to say/Decline	0 (0.0%)
		Unavailable	23 (4.3%)
Intra-axial hemorrhage	≤1.5mm		N=702
		Hispanic	57 (8.1%)
		Not Hispanic	602 (85.8%)
		Prefer not to say/Decline	3 (0.4%)
		Unavailable	40 (5.7%)
	>1.5mm & ≤5.0mm		N=875
		Hispanic	68 (7.8%)
		Not Hispanic	757 (86.5%)
		Prefer not to say/Decline	4 (0.5%)
		Unavailable	46 (5.3%)

Finding	Slice thickness range	Race	Total
Intraventricular hemorrhage	≤1.5mm		N=150
		Hispanic	17 (11.3%)
		Not Hispanic	129 (86.0%)
		Prefer not to say/Decline	1 (0.7%)
		Unavailable	3 (2.0%)
	>1.5mm & ≤5.0mm		N=184
		Hispanic	16 (8.7%)
		Not Hispanic	162 (88.0%)
		Prefer not to say/Decline	1 (0.5%)
		Unavailable	5 (2.7%)
Mass effect	≤1.5mm		N=626
		Hispanic	50 (8.0%)
		Not Hispanic	533 (85.1%)
		Prefer not to say/Decline	3 (0.5%)
		Unavailable	40 (6.4%)
	>1.5mm & ≤5.0mm		N=762
		Hispanic	51 (6.7%)
		Not Hispanic	669 (87.8%)
		Prefer not to say/Decline	4 (0.5%)
		Unavailable	38 (5.0%)
Obstructive hydrocephalus	≤1.5mm		N=175
		Hispanic	17 (9.7%)
		Not Hispanic	142 (81.1%)
		Prefer not to say/Decline	0 (0.0%)
		Unavailable	16 (9.1%)
	>1.5mm & ≤5.0mm		N=191
		Hispanic	17 (8.9%)
		Not Hispanic	157 (82.2%)
		Prefer not to say/Decline	0 (0.0%)
		Unavailable	17 (8.9%)

Finding	Slice thickness range	Race	Total
Vasogenic edema	≤1.5mm		N=167
		Hispanic	10 (6.0%)
		Not Hispanic	151 (90.4%)
		Prefer not to say/Decline	0 (0.0%)
		Unavailable	6 (3.6%)
	>1.5mm & ≤5.0mm		N=293
		Hispanic	20 (6.8%)
		Not Hispanic	264 (90.1%)
		Prefer not to say/Decline	1 (0.3%)
		Unavailable	8 (2.7%)

Equipment manufacturer

Results

The following tables outline the sensitivity and specificity for each finding by operating point and equipment manufacturer.

See:

- [CXR: Equipment manufacturer – sensitivity and specificity, page 35](#)
- [CTB: Equipment manufacturer – sensitivity and specificity, page 43](#)
- [CTB: Equipment manufacturer – Outside US \(OUS\) data analysis, page 48](#)

Additional information

The following additional information relates to the findings listed below:

Finding/s	Additional information
Mass effect and vasogenic edema	<ul style="list-style-type: none"> • Standalone performance evaluation of the device was performed on datasets with the majority of subgroup representation of studies acquired using GE Healthcare, Siemens and Toshiba CT scanners • Use of this device with other CT scanner manufacturers may result in differences in performance
Acute subarachnoid hemorrhage, acute subdural/epidural hematoma, intra-axial hemorrhage and intraventricular hemorrhage	<ul style="list-style-type: none"> • Standalone performance evaluation of the device was performed on datasets with the majority of subgroup representation of studies acquired using GE Healthcare, Siemens and Toshiba CT scanners • Use of this device with other CT scanner manufacturers may result in differences in performance • Additional analysis was conducted with GE, Philips, Siemens and Toshiba CT scanners on a dataset $\leq 1.5\text{mm}$ slice thickness sourced from Australia
Obstructive hydrocephalus	<ul style="list-style-type: none"> • Standalone performance evaluation of the device was performed on datasets with the majority of subgroup representation of studies acquired using GE Healthcare and Siemens CT scanners • Use of this device with other CT scanner manufacturers may result in differences in performance • Additional analysis was conducted with GE, Philips, Siemens and Toshiba CT scanners on a dataset $\leq 1.5\text{mm}$ slice thickness sourced from Australia

CXR: Equipment manufacturer – sensitivity and specificity

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion	0.380	Agfa	96.4 (92.9,99.1)	88.1 (81.2,94.1)
		Carestream	100.0 (100.0,100.0)	90.2 (80.5,97.6)
		Fujifilm	96.1 (90.8,100.0)	91.5 (85.4,97.6)
		GE Healthcare	100.0 (100.0,100.0)	94.7 (84.2,100.0)
		Kodak	100.0 (100.0,100.0)	91.7 (79.2,100.0)
		Konica Minolta	95.2 (90.5,99.0)	78.7 (69.3,88.0)
		McKesson	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	78.0 (66.1,88.1)
		Siemens	96.2 (88.5,100.0)	92.5 (84.9,98.1)
		Varian	84.6 (61.5,100.0)	76.9 (53.8,100.0)
		Multiple	97.1 (91.2,100.0)	78.3 (60.9,91.3)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.425	Agfa	90.2 (84.8,95.5)	94.1 (89.1,98.0)
		Carestream	100.0 (100.0,100.0)	92.7 (82.9,100.0)
		Fujifilm	94.7 (89.5,98.7)	96.3 (91.5,100.0)
		GE Healthcare	100.0 (100.0,100.0)	94.7 (84.2,100.0)
		Kodak	100.0 (100.0,100.0)	91.7 (79.2,100.0)
		Konica Minolta	93.3 (88.6,97.1)	80.0 (70.7,89.3)
		McKesson	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	79.7 (69.5,89.8)
		Siemens	96.2 (88.5,100.0)	92.5 (84.9,98.1)
		Varian	84.6 (61.5,100.0)	76.9 (53.8,100.0)
		Multiple	94.1 (85.3,100.0)	82.6 (65.2,95.7)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Pleural effusion (cont).	0.450	Agfa	90.2 (84.8,95.5)	94.1 (89.1,98.0)
		Carestream	100.0 (100.0,100.0)	92.7 (82.9,100.0)
		Fujifilm	94.7 (89.5,98.7)	98.8 (96.3,100.0)
		GE Healthcare	100.0 (100.0,100.0)	94.7 (84.2,100.0)
		Kodak	100.0 (100.0,100.0)	91.7 (79.2,100.0)
		Konica Minolta	89.5 (82.9,95.2)	84.0 (76.0,92.0)
		McKesson	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	98.3 (94.8,100.0)	83.1 (72.9,91.5)
		Siemens	96.2 (88.5,100.0)	92.5 (84.9,98.1)
		Varian	84.6 (61.5,100.0)	84.6 (61.5,100.0)
		Multiple	91.2 (79.4,100.0)	87.0 (73.9,100.0)
		Unknown	91.7 (75.0,100.0)	100.0 (100.0,100.0)
	0.475	Agfa	87.5 (81.2,92.9)	98.0 (95.0,100.0)
		Carestream	88.9 (77.8,100.0)	97.6 (92.7,100.0)
		Fujifilm	93.4 (88.2,98.7)	98.8 (96.3,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	100.0 (100.0,100.0)	91.7 (79.2,100.0)
		Konica Minolta	85.7 (79.0,92.4)	88.0 (80.0,94.7)
		McKesson	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	98.3 (94.8,100.0)	83.1 (72.9,91.5)
		Siemens	88.5 (73.1,100.0)	94.3 (88.7,100.0)
		Varian	84.6 (61.5,100.0)	84.6 (61.5,100.0)
		Multiple	88.2 (76.5,97.1)	91.3 (78.3,100.0)
		Unknown	83.3 (58.3,100.0)	100.0 (100.0,100.0)
	0.500	Agfa	87.5 (81.2,92.9)	99.0 (97.0,100.0)
		Carestream	85.2 (70.4,96.3)	97.6 (92.7,100.0)
		Fujifilm	90.8 (84.2,97.4)	100.0 (100.0,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	100.0 (100.0,100.0)	95.8 (87.5,100.0)
		Konica Minolta	83.8 (76.2,90.5)	90.7 (84.0,97.3)
		McKesson	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	96.6 (91.4,100.0)	88.1 (79.7,94.9)
		Siemens	80.8 (65.4,96.2)	94.3 (88.7,100.0)
		Varian	84.6 (61.5,100.0)	84.6 (61.5,100.0)
		Multiple	79.4 (64.7,91.2)	95.7 (87.0,100.0)
		Unknown	83.3 (58.3,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Pneumoperitoneum	0.250	Agfa	94.4 (86.1,100.0)	91.3 (84.1,97.1)
		Carestream	96.8 (90.3,100.0)	84.6 (73.1,94.2)
		Fujifilm	92.9 (78.6,100.0)	82.8 (69.0,96.6)
		Konica Minolta	100.0 (100.0,100.0)	93.3 (80.0,100.0)
		Philips	100.0 (100.0,100.0)	83.3 (58.3,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Multiple	100.0 (100.0,100.0)	66.7 (0.0,100.0)
	0.300	Agfa	91.7 (83.3,100.0)	92.8 (87.0,98.6)
		Carestream	93.5 (83.9,100.0)	88.5 (78.8,96.2)
		Fujifilm	92.9 (78.6,100.0)	82.8 (69.0,96.6)
		Konica Minolta	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	83.3 (58.3,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Multiple	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.350	Agfa	88.9 (77.8,97.2)	95.7 (89.9,100.0)
		Carestream	90.3 (80.6,100.0)	92.3 (84.6,98.1)
		Fujifilm	92.9 (78.6,100.0)	89.7 (79.3,100.0)
		Konica Minolta	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	83.3 (58.3,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Multiple	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.400	Agfa	86.1 (75.0,97.2)	97.1 (92.8,100.0)
		Carestream	90.3 (80.6,100.0)	94.2 (86.5,100.0)
		Fujifilm	92.9 (78.6,100.0)	89.7 (79.3,100.0)
		Konica Minolta	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Multiple	100.0 (100.0,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Pneumoperitoneum (cont.)	0.450	Agfa	80.6 (66.7,91.7)	98.6 (95.7,100.0)
		Carestream	87.1 (74.2,96.8)	98.1 (94.2,100.0)
		Fujifilm	92.9 (78.6,100.0)	96.6 (89.7,100.0)
		Konica Minolta	92.3 (76.9,100.0)	100.0 (100.0,100.0)
		Philips	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Multiple	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Pneumothorax	0.200	Agfa	96.2 (93.1,98.7)	80.6 (75.0,86.2)
		Carestream	97.3 (91.9,100.0)	87.0 (76.1,95.7)
		Fujifilm	100.0 (100.0,100.0)	94.3 (88.6,98.6)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	83.3 (50.0,100.0)	100.0 (100.0,100.0)
		Konica Minolta	98.4 (95.2,100.0)	83.3 (74.4,91.0)
		McKesson	100.0 (100.0,100.0)	92.3 (76.9,100.0)
		Philips	93.6 (85.1,100.0)	98.1 (94.2,100.0)
		Siemens	100.0 (100.0,100.0)	96.6 (89.7,100.0)
		Varian	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.250	Agfa	96.2 (93.1,98.7)	85.2 (80.1,89.8)
		Carestream	94.6 (86.5,100.0)	91.3 (82.6,97.8)
		Fujifilm	100.0 (100.0,100.0)	98.6 (95.7,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	83.3 (50.0,100.0)	100.0 (100.0,100.0)
		Konica Minolta	98.4 (95.2,100.0)	88.5 (80.8,94.9)
		McKesson	100.0 (100.0,100.0)	92.3 (76.9,100.0)
		Philips	91.5 (83.0,97.9)	100.0 (100.0,100.0)
		Siemens	92.9 (78.6,100.0)	100.0 (100.0,100.0)
		Varian	95.5 (86.4,100.0)	100.0 (100.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Pneumothorax (cont.)	0.300	Agfa	95.6 (92.5,98.7)	90.8 (86.7,94.4)
		Carestream	94.6 (86.5,100.0)	91.3 (82.6,97.8)
		Fujifilm	96.8 (90.3,100.0)	98.6 (95.7,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	83.3 (50.0,100.0)	100.0 (100.0,100.0)
		Konica Minolta	95.2 (88.7,100.0)	89.7 (83.3,96.2)
		McKesson	100.0 (100.0,100.0)	92.3 (76.9,100.0)
		Philips	91.5 (83.0,97.9)	100.0 (100.0,100.0)
		Siemens	92.9 (78.6,100.0)	100.0 (100.0,100.0)
		Varian	90.9 (77.3,100.0)	100.0 (100.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.350	Agfa	91.8 (87.4,95.6)	94.4 (90.8,97.4)
		Carestream	94.6 (86.5,100.0)	93.5 (84.8,100.0)
		Fujifilm	96.8 (90.3,100.0)	98.6 (95.7,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	66.7 (33.3,100.0)	100.0 (100.0,100.0)
		Konica Minolta	95.2 (88.7,100.0)	89.7 (83.3,96.2)
		McKesson	97.1 (91.4,100.0)	92.3 (76.9,100.0)
		Philips	91.5 (83.0,97.9)	100.0 (100.0,100.0)
		Siemens	92.9 (78.6,100.0)	100.0 (100.0,100.0)
		Varian	90.9 (77.3,100.0)	100.0 (100.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.400	Agfa	89.9 (84.9,94.3)	96.4 (93.9,99.0)
		Carestream	94.6 (86.5,100.0)	95.7 (89.1,100.0)
		Fujifilm	90.3 (77.4,100.0)	98.6 (95.7,100.0)
		GE Healthcare	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Kodak	66.7 (33.3,100.0)	100.0 (100.0,100.0)
		Konica Minolta	91.9 (83.9,98.4)	91.0 (84.6,97.4)
		McKesson	97.1 (91.4,100.0)	92.3 (76.9,100.0)
		Philips	87.2 (76.6,95.7)	100.0 (100.0,100.0)
		Siemens	92.9 (78.6,100.0)	100.0 (100.0,100.0)
		Varian	86.4 (68.2,100.0)	100.0 (100.0,100.0)
		Unknown	100.0 (100.0,100.0)	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Tension pneumothorax	0.225	Agfa	94.3 (88.6,98.6)	90.9 (87.4,94.0)
		Carestream	100.0 (100.0,100.0)	96.2 (91.2,100.0)
		Fujifilm	85.7 (57.1,100.0)	95.7 (91.5,98.9)
		GE Healthcare	-	90.0 (66.7,100.0)
		Kodak	-	100.0 (100.0,100.0)
		Konica Minolta	100.0 (100.0,100.0)	95.7 (91.4,99.1)
		McKesson	100.0 (100.0,100.0)	92.7 (82.9,100.0)
		Philips	100.0 (100.0,100.0)	98.9 (96.8,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	83.7 (72.1,95.3)
		Unknown	-	100.0 (100.0,100.0)
	0.250	Agfa	92.9 (85.7,98.6)	91.6 (88.1,94.7)
		Carestream	100.0 (100.0,100.0)	96.2 (91.2,100.0)
		Fujifilm	85.7 (57.1,100.0)	95.7 (91.5,98.9)
		GE Healthcare	-	90.0 (66.7,100.0)
		Kodak	-	100.0 (100.0,100.0)
		Konica Minolta	100.0 (100.0,100.0)	95.7 (91.4,99.1)
		McKesson	100.0 (100.0,100.0)	97.6 (92.7,100.0)
		Philips	100.0 (100.0,100.0)	98.9 (96.8,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	100.0 (100.0,100.0)	86.0 (74.4,95.3)
		Unknown	-	100.0 (100.0,100.0)
	0.300	Agfa	92.9 (85.7,98.6)	93.0 (89.9,95.8)
		Carestream	100.0 (100.0,100.0)	98.8 (96.2,100.0)
		Fujifilm	85.7 (57.1,100.0)	96.8 (92.6,100.0)
		GE Healthcare	-	90.0 (66.7,100.0)
		Kodak	-	100.0 (100.0,100.0)
		Konica Minolta	100.0 (100.0,100.0)	95.7 (91.4,99.1)
		McKesson	100.0 (100.0,100.0)	97.6 (92.7,100.0)
		Philips	80.0 (40.0,100.0)	98.9 (96.8,100.0)
		Siemens	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Varian	80.0 (40.0,100.0)	88.4 (79.1,97.7)
		Unknown	-	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Tension pneumothorax (cont.)	0.350	Agfa	88.6 (81.4,95.7)	94.7 (91.9,97.2)
		Carestream	100.0 (100.0,100.0)	98.8 (96.2,100.0)
		Fujifilm	85.7 (57.1,100.0)	96.8 (92.6,100.0)
		GE Healthcare	-	100.0 (100.0,100.0)
		Kodak	-	100.0 (100.0,100.0)
		Konica Minolta	100.0 (100.0,100.0)	96.6 (93.1,99.1)
		McKesson	100.0 (100.0,100.0)	97.6 (92.7,100.0)
		Philips	60.0 (20.0,100.0)	98.9 (96.8,100.0)
		Siemens	75.0 (25.0,100.0)	100.0 (100.0,100.0)
		Varian	80.0 (40.0,100.0)	93.0 (83.7,100.0)
		Unknown	-	100.0 (100.0,100.0)
	0.400	Agfa	85.7 (77.1,92.9)	96.5 (94.4,98.6)
		Carestream	100.0 (100.0,100.0)	98.8 (96.2,100.0)
		Fujifilm	85.7 (57.1,100.0)	97.9 (94.7,100.0)
		GE Healthcare	-	100.0 (100.0,100.0)
		Kodak	-	100.0 (100.0,100.0)
		Konica Minolta	100.0 (100.0,100.0)	97.4 (94.0,100.0)
		McKesson	100.0 (100.0,100.0)	97.6 (92.7,100.0)
		Philips	60.0 (20.0,100.0)	98.9 (96.8,100.0)
		Siemens	50.0 (0.0,100.0)	100.0 (100.0,100.0)
		Varian	80.0 (40.0,100.0)	93.0 (83.7,100.0)
		Unknown	-	100.0 (100.0,100.0)

Finding	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Vertebral compression fracture	0.460	Agfa	81.0 (61.9,95.2)	79.3 (65.5,93.1)
		Carestream	100.0 (100.0,100.0)	80.0 (60.0,95.0)
		Fujifilm	94.2 (88.4,98.6)	83.6 (74.0,91.8)
		GE Healthcare	75.0 (25.0,100.0)	100.0 (100.0,100.0)
		Kodak	94.7 (84.2,100.0)	84.2 (68.4,100.0)
		Philips	94.6 (86.5,100.0)	89.4 (80.9,97.9)
		Siemens	92.0 (80.0,100.0)	94.9 (87.2,100.0)
		Varian	91.9 (83.8,100.0)	90.7 (81.4,97.7)
		Multiple	100.0 (100.0,100.0)	79.2 (62.5,95.8)
		Unknown	94.7 (84.2,100.0)	80.0 (60.0,95.0)
	0.500	Agfa	81.0 (61.9,95.2)	86.2 (72.4,96.6)
		Carestream	100.0 (100.0,100.0)	85.0 (65.0,100.0)
		Fujifilm	92.8 (86.9,98.6)	90.4 (82.2,97.3)
		GE Healthcare	75.0 (25.0,100.0)	100.0 (100.0,100.0)
		Kodak	94.7 (84.2,100.0)	94.7 (84.2,100.0)
		Philips	94.6 (86.5,100.0)	93.6 (85.1,100.0)
		Siemens	92.0 (80.0,100.0)	97.4 (92.3,100.0)
		Varian	89.2 (78.4,97.3)	93.0 (83.7,100.0)
		Multiple	100.0 (100.0,100.0)	87.5 (75.0,100.0)
		Unknown	94.7 (84.2,100.0)	80.0 (60.0,95.0)
	0.550	Agfa	81.0 (61.9,95.2)	93.1 (82.8,100.0)
		Carestream	100.0 (100.0,100.0)	90.0 (75.0,100.0)
		Fujifilm	82.6 (73.9,91.3)	94.5 (89.0,98.6)
		GE Healthcare	75.0 (25.0,100.0)	100.0 (100.0,100.0)
		Kodak	94.7 (84.2,100.0)	100.0 (100.0,100.0)
		Philips	86.5 (73.0,97.3)	95.7 (89.4,100.0)
		Siemens	84.0 (68.0,96.0)	97.4 (92.3,100.0)
		Varian	81.1 (67.6,91.9)	95.3 (88.4,100.0)
		Multiple	100.0 (100.0,100.0)	95.8 (87.5,100.0)
		Unknown	94.7 (84.2,100.0)	85.0 (70.0,100.0)

CTB: Equipment manufacturer – sensitivity and specificity

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	≤1.5mm	0.014372	GE Healthcare	100.0 (100.0,100.0)	94.1 (85.3,100.0)
			NeuroLogica	-	100.0 (100.0,100.0)
			Siemens	96.8 (92.5,100.0)	86.0 (76.0,94.0)
		0.060162	GE Healthcare	94.4 (87.0,100.0)	94.1 (85.3,100.0)
			NeuroLogica	-	100.0 (100.0,100.0)
			Siemens	93.5 (88.2,97.8)	98.0 (94.0,100.0)
		0.082652	GE Healthcare	90.7 (81.5,98.1)	100.0 (100.0,100.0)
			NeuroLogica	-	100.0 (100.0,100.0)
			Siemens	89.2 (82.8,95.7)	100.0 (100.0,100.0)
	>1.5mm & ≤5.0mm	0.020255	GE Healthcare	91.8 (83.7,98.0)	90.9 (81.8,100.0)
			NeuroLogica	-	100.0 (100.0,100.0)
			Siemens	88.3 (81.9,94.7)	92.0 (84.0,98.0)
			Toshiba	94.9 (87.2,100.0)	95.2 (85.7,100.0)
		0.030010	GE Healthcare	89.8 (81.6,98.0)	97.0 (90.9,100.0)
			NeuroLogica	-	100.0 (100.0,100.0)
			Siemens	84.0 (76.6,91.5)	96.0 (90.0,100.0)
			Toshiba	92.3 (82.1,100.0)	95.2 (85.7,100.0)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subdural/epidural hematoma	≤1.5mm	0.060177	GE Healthcare	95.0 (91.0,99.0)	87.5 (75.0,96.9)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	89.5 (85.0,93.5)	86.2 (76.9,93.8)
			Toshiba	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		0.101143	GE Healthcare	93.0 (88.0,97.0)	96.9 (90.6,100.0)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	87.0 (82.0,91.5)	93.8 (87.7,98.5)
			Toshiba	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		0.135700	GE Healthcare	93.0 (88.0,97.0)	100.0 (100.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	83.0 (78.0,88.0)	95.4 (89.2,100.0)
			Toshiba	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	>1.5mm & ≤5.0mm	0.060177	GE Healthcare	84.8 (78.6,91.1)	90.5 (81.0,97.6)
			NeuroLogica	100.0 (100.0,100.0)	0.0 (0.0,0.0)
			Philips	100.0 (100.0,100.0)	-
			Siemens	81.2 (75.6,86.8)	93.8 (87.5,98.4)
			Toshiba	81.2 (72.9,89.4)	82.1 (67.9,96.4)
Intra-axial hemorrhage	≤1.5mm	0.322700	GE Healthcare	94.3 (89.6,98.1)	84.7 (74.6,93.2)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	92.7 (90.0,95.4)	85.8 (80.2,91.4)
			Toshiba	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	>1.5mm & ≤5.0mm	0.203600	GE Healthcare	93.5 (89.7,96.7)	83.8 (76.8,90.9)
			NeuroLogica	50.0 (0.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	93.2 (90.4,95.6)	86.1 (80.4,91.1)
			Toshiba	97.1 (88.2,100.0)	83.9 (71.0,96.8)
		0.322700	GE Healthcare	87.0 (82.1,91.8)	90.9 (84.8,96.0)
			NeuroLogica	50.0 (0.0,100.0)	-
			Philips	100.0 (100.0,100.0)	-
			Siemens	91.8 (89.1,94.3)	89.9 (85.4,94.3)
			Toshiba	94.1 (85.3,100.0)	90.3 (80.6,100.0)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Intraventricular hemorrhage	≤1.5mm	0.015487	GE Healthcare	95.8 (87.5,100.0)	91.7 (79.2,100.0)
			NeuroLogica	100.0 (100.0,100.0)	100.0 (100.0,100.0)
			Siemens	95.8 (89.6,100.0)	90.4 (82.7,98.1)
		0.051859	GE Healthcare	91.7 (79.2,100.0)	95.8 (87.5,100.0)
			NeuroLogica	0.0 (0.0,0.0)	100.0 (100.0,100.0)
			Siemens	91.7 (83.3,97.9)	98.1 (94.2,100.0)
	>1.5mm & ≤5.0mm	0.008430	GE Healthcare	91.7 (79.2,100.0)	95.7 (87.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	100.0 (100.0,100.0)
			Siemens	95.8 (89.6,100.0)	79.2 (66.7,89.6)
			Toshiba	100.0 (100.0,100.0)	90.5 (76.2,100.0)
		0.015487	GE Healthcare	87.5 (75.0,100.0)	95.7 (87.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	100.0 (100.0,100.0)
			Siemens	91.7 (83.3,97.9)	85.4 (75.0,93.8)
			Toshiba	100.0 (100.0,100.0)	90.5 (76.2,100.0)
		0.051859	GE Healthcare	83.3 (66.7,95.8)	100.0 (100.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	100.0 (100.0,100.0)
			Siemens	85.4 (75.0,95.8)	97.9 (93.8,100.0)
			Toshiba	100.0 (100.0,100.0)	95.2 (85.7,100.0)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Mass effect	≤1.5mm	0.160195	GE Healthcare	97.1 (94.3,99.4)	83.7 (72.1,93.0)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	83.3 (50.0,100.0)	-
			Siemens	97.4 (95.4,99.0)	91.1 (84.4,96.7)
			Toshiba	75.0 (0.0,100.0)	-
		0.221484	GE Healthcare	96.0 (93.1,98.3)	86.0 (74.4,95.3)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	83.3 (50.0,100.0)	-
			Siemens	97.4 (95.4,99.0)	91.1 (84.4,96.7)
			Toshiba	75.0 (0.0,100.0)	-
	>1.5mm & ≤5.0mm	0.120944	GE Healthcare	96.6 (93.7,98.9)	82.5 (70.0,92.5)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	83.3 (42.9,100.0)	-
			Siemens	97.2 (95.1,99.0)	92.0 (85.1,96.6)
			Toshiba	96.7 (93.4,99.2)	90.2 (80.5,97.6)
		0.160195	GE Healthcare	94.9 (91.4,97.7)	90.0 (80.0,97.5)
			NeuroLogica	100.0 (100.0,100.0)	-
			Philips	83.3 (42.9,100.0)	-
			Siemens	95.5 (93.1,97.6)	94.3 (88.5,98.9)
			Toshiba	95.9 (91.8,99.2)	92.7 (85.3,100.0)
Obstructive hydrocephalus	≤1.5mm	0.149943	GE Healthcare	100.0 (100.0,100.0)	84.2 (68.4,100.0)
			Siemens	96.9 (92.3,100.0)	96.2 (91.0,100.0)
		0.185900	GE Healthcare	100.0 (100.0,100.0)	84.2 (68.4,100.0)
			Siemens	93.8 (87.7,98.5)	97.4 (93.6,100.0)
		0.281473	GE Healthcare	100.0 (100.0,100.0)	94.7 (84.2,100.0)
			Siemens	90.8 (83.1,96.9)	97.4 (93.6,100.0)
	>1.5mm & ≤5.0mm	0.100591	GE Healthcare	100.0 (100.0,100.0)	87.5 (75.0,100.0)
			Siemens	96.9 (92.3,100.0)	97.5 (93.7,100.0)
		0.149943	GE Healthcare	100.0 (100.0,100.0)	87.5 (75.0,100.0)
			Siemens	95.4 (89.2,100.0)	97.5 (93.7,100.0)
		0.185900	GE Healthcare	100.0 (100.0,100.0)	87.5 (75.0,100.0)
			Siemens	93.8 (87.7,98.5)	97.5 (93.7,100.0)
		0.281473	GE Healthcare	93.3 (80.0,100.0)	87.5 (75.0,100.0)
			Siemens	87.7 (80.0,95.4)	97.5 (93.7,100.0)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Vasogenic edema	≤1.5mm	0.060584	GE Healthcare	71.4 (42.9,100.0)	91.7 (75.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	50.0 (0.0,100.0)
			Siemens	94.2 (86.5,100.0)	91.2 (84.6,96.7)
		0.094076	GE Healthcare	71.4 (42.9,100.0)	91.7 (75.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	75.0 (25.0,100.0)
			Siemens	92.3 (84.6,98.1)	91.2 (84.6,96.7)
		0.145352	GE Healthcare	71.4 (42.9,100.0)	91.7 (75.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	100.0 (100.0,100.0)
			Siemens	92.3 (84.6,98.1)	93.4 (87.9,97.8)
	>1.5mm & ≤5.0mm	0.060584	GE Healthcare	89.3 (78.6,100.0)	98.0 (94.1,100.0)
			NeuroLogica	100.0 (100.0,100.0)	60.0 (20.0,100.0)
			Siemens	96.2 (90.6,100.0)	90.5 (84.2,95.8)
			Toshiba	100.0 (100.0,100.0)	97.7 (93.0,100.0)
		0.094076	GE Healthcare	89.3 (78.6,100.0)	98.0 (94.1,100.0)
			NeuroLogica	100.0 (100.0,100.0)	80.0 (40.0,100.0)
			Siemens	94.3 (86.8,100.0)	91.6 (85.3,96.8)
			Toshiba	100.0 (100.0,100.0)	97.7 (93.0,100.0)
		0.145352	GE Healthcare	89.3 (78.6,100.0)	98.0 (94.1,100.0)
			NeuroLogica	100.0 (100.0,100.0)	80.0 (40.0,100.0)
			Siemens	90.6 (81.1,98.1)	92.6 (87.4,97.9)
			Toshiba	94.1 (82.4,100.0)	100.0 (100.0,100.0)
		0.261255	GE Healthcare	85.7 (71.4,96.4)	100.0 (100.0,100.0)
			NeuroLogica	100.0 (100.0,100.0)	80.0 (40.0,100.0)
			Siemens	90.6 (81.1,98.1)	95.8 (91.6,98.9)
			Toshiba	94.1 (82.4,100.0)	100.0 (100.0,100.0)

CTB: Equipment manufacturer – Outside US (OUS) data analysis

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	≤1.5mm	0.014372	GE Healthcare	97.5 (91.7, 100.0)	76.6 (73.4, 79.8)
			Phillips	90.9 (76.9, 100.0)	79.8 (75.7, 83.7)
			Siemens	100.0 (100.0, 100.0)	68.3 (63.5, 73.0)
			Toshiba	98.2 (93.9, 100.0)	81.3 (79.1, 83.5)
		0.060162	GE Healthcare	90.0 (79.5, 97.7)	88.9 (86.4, 91.2)
			Phillips	90.9 (76.9, 100.0)	92.6 (89.8, 95.1)
			Siemens	97.0 (90.0, 100.0)	85.3 (81.6, 88.8)
			Toshiba	92.8 (85.2, 98.4)	92.4 (90.9, 93.9)
		0.082652	GE Healthcare	90.0 (79.5, 97.7)	90.8 (88.5, 92.9)
			Phillips	90.9 (76.9, 100.0)	94.1 (91.6, 96.3)
			Siemens	97.0 (90.0, 100.0)	88.3 (85.0, 91.5)
			Toshiba	92.8 (85.2, 98.4)	93.6 (92.1, 95.0)
Acute subdural/epidural hematoma	≤1.5mm	0.030172	GE Healthcare	96.1 (89.8, 100.0)	84.7 (82.0, 87.3)
			Philips	100.0 (100.0, 100.0)	89.7 (86.5, 92.6)
			Siemens	95.9 (89.5, 100.0)	79.1 (74.9, 83.3)
			Toshiba	96.1 (89.8, 100.0)	84.7 (82.0, 87.3)
		0.060177	GE Healthcare	90.2 (81.1, 97.8)	88.7 (86.3, 91.0)
			Philips	100.0 (100.0, 100.0)	92.8 (90.1, 95.3)
			Siemens	93.8 (86.3, 100.0)	85.9 (82.2, 89.5)
			Toshiba	94.9 (88.5, 100.0)	93.1 (91.7, 94.6)
		0.101143	GE Healthcare	88.3 (78.6, 96.2)	92.0 (89.9, 94.0)
			Philips	96.2 (87.0, 100.0)	94.1 (91.6, 96.3)
			Siemens	89.6 (80.0, 97.7)	89.0 (85.7, 92.1)
			Toshiba	93.2 (86.1, 98.5)	94.6 (93.3, 95.9)
		0.135700	GE Healthcare	88.3 (78.6, 96.2)	93.6 (91.7, 95.4)
			Philips	96.2 (87.0, 100.0)	95.9 (93.8, 97.7)
			Siemens	89.6 (80.0, 97.7)	90.7 (87.6, 93.6)
			Toshiba	91.5 (83.6, 98.1)	95.2 (94.0, 96.4)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Intra-axial hemorrhage	≤1.5mm	0.1498	GE Healthcare	85.9 (78.0, 92.9)	92.6 (90.5, 94.7)
			Philips	78.1 (62.5, 92.0)	89.5 (86.3, 92.4)
			Siemens	93.0 (85.7, 98.4)	89.6 (86.2, 92.7)
			Toshiba	87.3 (81.3, 92.7)	93.5 (92.1, 94.9)
		0.2036	GE Healthcare	84.7 (76.7, 92.0)	94.8 (93.0, 96.5)
			Philips	78.1 (62.5, 92.0)	90.8 (87.7, 93.5)
			Siemens	91.2 (83.1, 98.1)	91.3 (88.2, 94.2)
			Toshiba	84.9 (78.5, 90.9)	94.9 (93.6, 96.1)
		0.322700	GE Healthcare	77.7 (68.4, 86.1)	96.7 (95.2, 98.0)
			Philips	68.6 (51.6, 84.6)	94.5 (92.1, 96.7)
			Siemens	84.2 (74.0, 93.2)	95.3 (92.9, 97.4)
			Toshiba	77.0 (69.4, 84.1)	97.1 (96.1, 98.0)
Intraventricular hemorrhage	≤1.5mm	0.00843	GE Healthcare	100.0 (100.0, 100.0)	78.3 (75.2, 81.3)
			Philips	94.0 (80.0, 100.0)	84.1 (80.4, 87.7)
			Siemens	100.0 (100.0, 100.0)	71.6 (67.1, 76.1)
			Toshiba	100.0 (100.0, 100.0)	83.1 (81.0, 85.2)
		0.015487	GE Healthcare	100.0 (100.0, 100.0)	86.8 (84.3, 89.3)
			Philips	88.2 (69.6, 100.0)	89.6 (86.5, 92.6)
			Siemens	100.0 (100.0, 100.0)	83.2 (79.4, 86.9)
			Toshiba	100.0 (100.0, 100.0)	88.0 (86.1, 89.8)
		0.051859	GE Healthcare	90.9 (80.0, 100.0)	94.4 (92.6, 96.0)
			Philips	88.2 (69.6, 100.0)	97.5 (95.8, 98.9)
			Siemens	100.0 (100.0, 100.0)	93.6 (91.0, 95.9)
			Toshiba	100.0 (100.0, 100.0)	95.0 (93.8, 96.2)

Finding	Slice thickness range	Operating point	Manufacturer	Sensitivity % (95% CI)	Specificity % (95% CI)
Obstructive hydrocephalus	≤1.5mm	0.100591	GE Healthcare	97.0 (90.0, 100.0)	90.3 (88.1, 92.4)
			Philips	88.4 (71.4, 100.0)	92.9 (90.3, 95.3)
			Siemens	100.0 (100.0, 100.0)	89.7 (86.5, 92.6)
			Toshiba	88.4 (75.0, 100.0)	91.7 (90.1, 93.2)
		0.149943	GE Healthcare	97.0 (90.0, 100.0)	94.0 (92.3, 95.8)
			Philips	82.6 (61.9, 100.0)	94.2 (91.8, 96.4)
			Siemens	87.4 (68.7, 100.0)	92.5 (89.8, 95.0)
			Toshiba	88.4 (75.0, 100.0)	93.7 (92.4, 95.0)
		0.185900	GE Healthcare	97.0 (90.0, 100.0)	95.2 (93.6, 96.7)
			Philips	82.6 (61.9, 100.0)	95.2 (93.0, 97.2)
			Siemens	87.4 (68.7, 100.0)	93.3 (90.7, 95.7)
			Toshiba	80.7 (64.0, 95.2)	95.0 (93.7, 96.1)
		0.281473	GE Healthcare	97.0 (90.0, 100.0)	97.2 (96.0, 98.4)
			Philips	70.8 (47.1, 92.3)	97.5 (95.8, 98.9)
			Siemens	74.8 (50.0, 94.4)	95.4 (93.2, 97.4)
			Toshiba	76.8 (59.3, 92.3)	96.7 (95.7, 97.7)

Series slice thickness for CTB findings

Results

The following table outlines the sensitivity and specificity for each CTB finding by operating point and >1.5mm to ≤ 5mm slice thickness range.

CTB: Series thickness for slice thickness range >1.5mm to ≤5mm

Finding	Operating point	Slice thickness range	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	0.020255	>1.5 to ≤3.5 mm series	98.2 (94.5,100.0)	93.5 (83.9,100.0)
		>3.5 to ≤5 mm series	87.4 (81.1,92.9)	91.9 (85.1,97.3)
	0.030010	>1.5 to ≤3.5 mm series	96.4 (90.9,100.0)	96.8 (90.3,100.0)
		>3.5 to ≤5 mm series	83.5 (77.2,89.8)	95.9 (90.5,100.0)
Acute subdural/epidural hematoma	0.060177	>1.5 to ≤3.5 mm series	93.1 (88.2,97.1)	89.7 (79.5,97.4)
		>3.5 to ≤5 mm series	78.6 (73.9,83.1)	89.6 (83.3,94.8)
Intra-axial hemorrhage	0.322700	>1.5 to ≤3.5 mm series	95.5 (92.6,97.9)	88.8 (81.6,94.9)
		>3.5 to ≤5 mm series	86.7 (82.9,90.1)	91.1 (86.8,94.7)
	0.203600	>1.5 to ≤3.5 mm series	97.9 (95.9,99.6)	81.6 (74.5,88.8)
		>3.5 to ≤5 mm series	90.1 (86.9,92.8)	86.8 (81.6,91.6)
Intraventricular hemorrhage	0.015487	>1.5 to ≤3.5 mm series	92.9 (82.1,100.0)	83.3 (66.7,95.8)
		>3.5 to ≤5 mm series	92.1 (85.7,98.4)	91.3 (84.1,97.1)
	0.051859	>1.5 to ≤3.5 mm series	92.9 (82.1,100.0)	95.8 (87.5,100.0)
		>3.5 to ≤5 mm series	85.7 (77.8,93.7)	98.6 (95.7,100.0)
	0.008430	>1.5 to ≤3.5 mm series	96.4 (89.3,100.0)	83.3 (66.7,95.8)
		>3.5 to ≤5 mm series	95.2 (88.9,100.0)	87.0 (78.3,94.2)
Mass effect	0.120944	>1.5 to ≤3.5 mm series	98.5 (96.4,100.0)	88.4 (79.1,97.7)
		>3.5 to ≤5 mm series	96.0 (94.0,98.0)	89.6 (84.0,94.4)
	0.160195	>1.5 to ≤3.5 mm series	97.9 (95.9,100.0)	90.7 (81.4,97.7)
		>3.5 to ≤5 mm series	94.0 (91.5,96.2)	93.6 (88.8,97.6)
Obstructive hydrocephalus	0.100591	>1.5 to ≤3.5 mm series	97.9 (93.8,100.0)	96.0 (90.0,100.0)
		>3.5 to ≤5 mm series	97.2 (91.7,100.0)	94.7 (89.4,100.0)
	0.149943	>1.5 to ≤3.5 mm series	95.8 (89.6,100.0)	96.0 (90.0,100.0)
		>3.5 to ≤5 mm series	94.4 (86.1,100.0)	94.7 (89.4,100.0)
	0.185900	>1.5 to ≤3.5 mm series	93.8 (85.4,100.0)	96.0 (90.0,100.0)
		>3.5 to ≤5 mm series	94.4 (86.1,100.0)	94.7 (89.4,100.0)
	0.281473	>1.5 to ≤3.5 mm series	85.4 (75.0,93.8)	96.0 (90.0,100.0)
		>3.5 to ≤5 mm series	91.7 (80.6,100.0)	94.7 (89.4,100.0)

Finding	Operating point	Slice thickness range	Sensitivity % (95% CI)	Specificity % (95% CI)
Vasogenic edema	0.060584	>1.5 to ≤3.5 mm series	95.0 (87.5,100.0)	90.9 (84.4,96.1)
		>3.5 to ≤5 mm series	94.9 (89.8,100.0)	94.9 (90.6,98.3)
	0.094076	>1.5 to ≤3.5 mm series	95.0 (87.5,100.0)	92.2 (85.7,97.4)
		>3.5 to ≤5 mm series	93.2 (86.4,98.3)	95.7 (91.5,99.1)
	0.145352	>1.5 to ≤3.5 mm series	95.0 (87.5,100.0)	93.5 (88.3,98.7)
		>3.5 to ≤5 mm series	88.1 (79.7,94.9)	96.6 (93.2,99.1)
	0.261255	>1.5 to ≤3.5 mm series	95.0 (87.5,100.0)	97.4 (93.5,100.0)
		>3.5 to ≤5 mm series	86.4 (77.9,94.9)	97.4 (94.0,100.0)

Presence of co-existing findings or abnormalities

Results

The following tables outline:

- the sensitivity and specificity for each finding by operating point and the presence of co-existing findings or abnormalities
- the presence of co-existing findings or abnormalities by slice thickness range for each finding*
- the sensitivity and specificity by operating point and the presence of co-existing findings or abnormalities for intra-axial hemorrhage

*Any finding' means that one or more of the co-existing findings or abnormalities listed in the table is present.

See:

- [CTB: Presence of co-existing findings or abnormalities – sensitivity and specificity, page 54](#)
- [CTB: Presence of co-existing findings or abnormalities – distribution, page 57](#)
- [CTB: Presence of specific co-existing findings or abnormalities for intra-axial hemorrhage, page 64](#)

CTB: Presence of co-existing findings or abnormalities – sensitivity and specificity

Finding	Operating point	Co-finding	Sensitivity % (95% CI)	Specificity % (95% CI)
Acute subarachnoid hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$			
	0.014372	Any finding	98.3 (95.7,100.0)	78.8 (63.6,90.9)
		No additional findings	96.8 (90.3,100.0)	96.2 (90.4,100.0)
	0.060162	Any finding	96.6 (93.1,99.1)	93.9 (84.8,100.0)
		No additional findings	83.9 (71.0,96.8)	98.1 (94.2,100.0)
	0.082652	Any finding	91.4 (86.2,96.6)	100.0 (100.0,100.0)
		No additional findings	83.9 (71.0,96.8)	100.0 (100.0,100.0)
	Slice thickness range: $>1.5\text{mm}$ & $\leq 5.0\text{mm}$			
	0.020255	Any finding	92.4 (87.6,96.6)	86.4 (75.0,95.5)
		No additional findings	83.8 (70.3,94.6)	96.7 (91.8,100.0)
	0.030010	Any finding	89.7 (84.1,94.5)	93.2 (84.1,100.0)
		No additional findings	78.4 (64.9,91.9)	98.4 (95.1,100.0)
Acute subdural/epidural hematoma	Slice thickness range: $\leq 1.5\text{mm}$			
	0.060177	Any finding	93.5 (89.8,96.8)	80.9 (70.2,91.5)
		No additional findings	86.2 (79.3,93.1)	92.2 (84.3,98.0)
	0.101143	Any finding	90.3 (86.6,94.0)	93.6 (85.1,100.0)
		No additional findings	86.2 (79.3,93.1)	96.1 (90.2,100.0)
	0.135700	Any finding	87.0 (82.4,91.2)	95.7 (89.4,100.0)
		No additional findings	85.1 (77.0,92.0)	98.0 (94.1,100.0)
	Slice thickness range: $>1.5\text{mm}$ & $\leq 5.0\text{mm}$			
	0.060177	Any finding	85.0 (80.6,89.0)	85.7 (76.2,93.7)
		No additional findings	76.6 (69.4,83.9)	93.1 (87.5,98.6)
Intra-axial hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$			
	0.322700	Any finding	93.7 (91.4,95.8)	77.5 (70.3,84.8)
		No additional findings	40.0 (0.0,80.0)	98.8 (96.4,100.0)
	Slice thickness range: $>1.5\text{mm}$ & $\leq 5.0\text{mm}$			
	0.203600	Any finding	93.9 (91.9,95.8)	76.5 (69.9,82.5)
		No additional findings	55.6 (22.2,88.9)	100.0 (100.0,100.0)
	0.322700	Any finding	91.3 (89.1,93.6)	84.7 (79.2,89.6)
		No additional findings	22.2 (0.0,55.6)	100.0 (100.0,100.0)

Finding	Operating point	Co-finding	Sensitivity % (95% CI)	Specificity % (95% CI)
Intraventricular hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$			
	0.015487	Any finding	95.8 (90.3,100.0)	83.3 (71.4,92.9)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.051859	Any finding	90.3 (83.3,95.8)	95.2 (88.1,100.0)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$			
	0.008430	Any finding	95.5 (91.0,98.9)	75.0 (63.5,86.5)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.015487	Any finding	92.1 (86.5,97.8)	80.8 (69.2,90.4)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.051859	Any finding	87.6 (80.9,94.4)	96.2 (90.4,100.0)
		No additional findings	100.0 (100.0,100.0)	100.0(100.0,100.0)
Mass effect	Slice thickness range: $\leq 1.5\text{mm}$			
	0.160195	Any finding	97.3 (95.8,98.5)	75.0 (62.5,85.7)
		No additional findings	88.2 (70.6,100.0)	98.7 (96.1,100.0)
	0.221484	Any finding	97.1 (95.4,98.5)	76.8 (66.1,87.5)
		No additional findings	82.4 (64.7,100.0)	98.7 (96.1,100.0)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$			
	0.120944	Any finding	97.2 (95.8,98.4)	76.8 (66.7,85.5)
		No additional findings	86.4 (68.2,100.0)	98.0 (94.9,100.0)
	0.160195	Any finding	95.8 (94.1,97.4)	82.6 (73.9,91.3)
		No additional findings	81.8 (63.6,95.5)	100.0 (100.0,100.0)

Finding	Operating point	Co-finding	Sensitivity % (95% CI)	Specificity % (95% CI)
Obstructive hydrocephalus	Slice thickness range: $\leq 1.5\text{mm}$			
	0.149943	Any finding	97.3 (93.3,100.0)	89.7 (81.0,96.6)
		No additional findings	-	100.0 (100.0,100.0)
	0.185900	Any finding	94.7 (89.3,98.7)	91.4 (82.8,98.3)
		No additional findings	-	100.0 (100.0,100.0)
	0.281473	Any finding	92.0 (85.3,97.3)	94.8 (87.9,100.0)
		No additional findings	-	100.0 (100.0,100.0)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$			
	0.100591	Any finding	97.6 (94.0,100.0)	91.7 (85.0,98.3)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.149943	Any finding	95.2 (90.4,98.8)	91.7 (85.0,98.3)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.185900	Any finding	94.0 (88.0,98.8)	91.7 (85.0,98.3)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.281473	Any finding	88.0 (80.7,94.0)	91.7 (85.0,98.3)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
Vasogenic edema	Slice thickness range: $\leq 1.5\text{mm}$			
	0.060584	Any finding	91.5 (84.7,98.3)	80.0 (68.0,90.0)
		No additional findings	100.0 (100.0,100.0)	98.2 (94.7,100.0)
	0.094076	Any finding	89.8 (81.4,96.6)	82.0 (72.0,92.0)
		No additional findings	100.0 (100.0,100.0)	98.2 (94.7,100.0)
	0.145352	Any finding	89.8 (81.4,96.6)	86.0 (76.0,94.0)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$			
	0.060584	Any finding	94.8 (89.7,99.0)	80.6 (70.1,89.6)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.094076	Any finding	93.8 (88.7,97.9)	83.6 (74.6,92.5)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.145352	Any finding	90.7 (84.5,95.9)	86.6 (77.6,94.0)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)
	0.261255	Any finding	89.7 (83.5,94.8)	92.5 (85.1,98.5)
		No additional findings	100.0 (100.0,100.0)	100.0 (100.0,100.0)

CTB: Presence of co-existing findings or abnormalities – distribution

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Acute subarachnoid hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$	(N = 147)	(N = 85)	(N = 232)
	Other intracranial hemorrhage	96 (65.3%)	5 (5.9%)	101 (43.5%)
	Fracture including calvarial/ skull base/facial	27 (18.4%)	1 (1.2%)	28 (12.1%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	70 (47.6%)	20 (23.5%)	90 (38.8%)
	Parenchymal atrophy (excluding age-expected atrophy)	6 (4.1%)	2 (2.4%)	8 (3.4%)
	Edema including transependymal/vasogenic	50 (34.0%)	6 (7.1%)	56 (24.1%)
	Hydrocephalus including obstructive/non-obstructive	17 (11.6%)	5 (5.9%)	22 (9.5%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	5 (3.4%)	5 (5.9%)	10 (4.3%)
	Other evidence of intracranial surgery	11 (7.5%)	4 (4.7%)	15 (6.5%)
	Metallic artifact overlapping region of interest	5 (3.4%)	5 (5.9%)	10 (4.3%)
	Motion artifact overlapping region of interest	9 (6.1%)	7 (8.2%)	16 (6.9%)
	Any finding	116 (78.9%)	33 (38.8%)	149 (64.2%)
	Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$	N=182	N=105	N=287
	Other intracranial hemorrhage	123 (67.6%)	6 (5.7%)	129 (44.9%)
	Fracture including calvarial/ skull base/facial	31 (17.0%)	1 (1.0%)	32 (11.1%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	82 (45.1%)	26 (24.8%)	108 (37.6%)
	Parenchymal atrophy (excluding age-expected atrophy)	9 (4.9%)	5 (4.8%)	14 (4.9%)
	Edema including transependymal/vasogenic	56 (30.8%)	7 (6.7%)	63 (22.0%)
	Hydrocephalus including obstructive/non-obstructive	18 (9.9%)	6 (5.7%)	24 (8.4%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	5 (2.7%)	5 (4.8%)	10 (3.5%)
	Other evidence of intracranial surgery	14 (7.7%)	7 (6.7%)	21 (7.3%)
	Metallic artifact overlapping region of interest	6 (3.3%)	7 (6.7%)	13 (4.5%)
	Motion artifact overlapping region of interest	11 (6.0%)	8 (7.6%)	19 (6.6%)
	Any finding	145 (79.7%)	44 (41.9%)	189 (65.9%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Acute subdural/epidural hematoma	Slice thickness range: $\leq 1.5\text{mm}$	N=303	N=98	N=401
	Other intracranial hemorrhage	131 (43.2%)	10 (10.2%)	141 (35.2%)
	Fracture including calvarial/ skull base/facial	56 (18.5%)	2 (2.0%)	58 (14.5%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	114 (37.6%)	32 (32.7%)	146 (36.4%)
	Parenchymal atrophy (excluding age-expected atrophy)	7 (2.3%)	7 (7.1%)	14 (3.5%)
	Edema including transependymal/vasogenic	72 (23.8%)	7 (7.1%)	79 (19.7%)
	Hydrocephalus including obstructive/non-obstructive	27 (8.9%)	6 (6.1%)	33 (8.2%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	15 (5.0%)	4 (4.1%)	19 (4.7%)
	Other evidence of intracranial surgery	53 (17.5%)	6 (6.1%)	59 (14.7%)
	Metallic artifact overlapping region of interest	11 (3.6%)	2 (2.0%)	13 (3.2%)
	Motion artifact overlapping region of interest	32 (10.6%)	9 (9.2%)	41 (10.2%)
	Any finding	216 (71.3%)	47 (48.0%)	263 (65.6%)
	Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$	N=397	N=135	N=532
	Other intracranial hemorrhage	171 (43.1%)	16 (11.9%)	187 (35.2%)
	Fracture including calvarial/ skull base/facial	74 (18.6%)	2 (1.5%)	76 (14.3%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	154 (38.8%)	45 (33.3%)	199 (37.4%)
	Parenchymal atrophy (excluding age-expected atrophy)	12 (3.0%)	7 (5.2%)	19 (3.6%)
	Edema including transependymal/vasogenic	96 (24.2%)	9 (6.7%)	105 (19.7%)
	Hydrocephalus including obstructive/non-obstructive	34 (8.6%)	6 (4.4%)	40 (7.5%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	17 (4.3%)	5 (3.7%)	22 (4.1%)
	Other evidence of intracranial surgery	65 (16.4%)	11 (8.1%)	76 (14.3%)
	Metallic artifact overlapping region of interest	13 (3.3%)	3 (2.2%)	16 (3.0%)
	Motion artifact overlapping region of interest	30 (7.6%)	11 (8.1%)	41 (7.7%)
	Any finding	273 (68.8%)	63 (46.7%)	336 (63.2%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Intra-axial hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$	N=480	N=222	N=702
	Other intracranial hemorrhage	311 (64.8%)	49 (22.1%)	360 (51.3%)
	Fracture including calvarial/ skull base/facial	40 (8.3%)	11 (5.0%)	51 (7.3%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	127 (26.5%)	55 (24.8%)	182 (25.9%)
	Parenchymal atrophy (excluding age-expected atrophy)	103 (21.5%)	60 (27.0%)	163 (23.2%)
	Edema including transependymal/vasogenic	435 (90.6%)	22 (9.9%)	457 (65.1%)
	Hydrocephalus including obstructive/non-obstructive	82 (17.1%)	11 (5.0%)	93 (13.2%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	26 (5.4%)	5 (2.3%)	31 (4.4%)
	Other evidence of intracranial surgery	71 (14.8%)	16 (7.2%)	87 (12.4%)
	Metallic artifact overlapping region of interest	39 (8.1%)	18 (8.1%)	57 (8.1%)
	Motion artifact overlapping region of interest	46 (9.6%)	14 (6.3%)	60 (8.5%)
	Any finding	475 (99.0%)	138 (62.2%)	613 (87.3%)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$	N=587	N=288	N=875
	Other intracranial hemorrhage	370 (63.0%)	71 (24.7%)	441 (50.4%)
	Fracture including calvarial/ skull base/facial	47 (8.0%)	16 (5.6%)	63 (7.2%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	153 (26.1%)	67 (23.3%)	220 (25.1%)
	Parenchymal atrophy (excluding age-expected atrophy)	134 (22.8%)	83 (28.8%)	217 (24.8%)
	Edema including transependymal/vasogenic	525 (89.4%)	24 (8.3%)	549 (62.7%)
	Hydrocephalus including obstructive/non-obstructive	96 (16.4%)	14 (4.9%)	110 (12.6%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	29 (4.9%)	6 (2.1%)	35 (4.0%)
	Other evidence of intracranial surgery	74 (12.6%)	17 (5.9%)	91 (10.4%)
	Metallic artifact overlapping region of interest	41 (7.0%)	16 (5.6%)	57 (6.5%)
	Motion artifact overlapping region of interest	55 (9.4%)	14 (4.9%)	69 (7.9%)
	Any finding	578 (98.5%)	183 (63.5%)	761 (87.0%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Intraventricular hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$	N=73	N=77	N=150
	Other intracranial hemorrhage	64 (87.7%)	11 (14.3%)	75 (50.0%)
	Fracture including calvarial/ skull base/facial	6 (8.2%)	2 (2.6%)	8 (5.3%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	64 (87.7%)	27 (35.1%)	91 (60.7%)
	Parenchymal atrophy (excluding age-expected atrophy)	1 (1.4%)	2 (2.6%)	3 (2.0%)
	Edema including transependymal/vasogenic	57 (78.1%)	11 (14.3%)	68 (45.3%)
	Hydrocephalus including obstructive/non-obstructive	31 (42.5%)	3 (3.9%)	34 (22.7%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	8 (11.0%)	2 (2.6%)	10 (6.7%)
	Other evidence of intracranial surgery	13 (17.8%)	11 (14.3%)	24 (16.0%)
	Metallic artifact overlapping region of interest	1 (1.4%)	2 (2.6%)	3 (2.0%)
	Motion artifact overlapping region of interest	4 (5.5%)	2 (2.6%)	6 (4.0%)
	Any finding	72 (98.6%)	42 (54.5%)	114 (76.0%)
	Slice thickness range: $> 1.5\text{mm}$ & $\leq 5.0\text{mm}$	N=91	N=93	N=184
	Other intracranial hemorrhage	80 (87.9%)	15 (16.1%)	95 (51.6%)
	Fracture including calvarial/ skull base/facial	6 (6.6%)	3 (3.2%)	9 (4.9%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	80 (87.9%)	33 (35.5%)	113 (61.4%)
	Parenchymal atrophy (excluding age-expected atrophy)	1 (1.1%)	4 (4.3%)	5 (2.7%)
	Edema including transependymal/vasogenic	71 (78.0%)	12 (12.9%)	83 (45.1%)
	Hydrocephalus including obstructive/non-obstructive	39 (42.9%)	3 (3.2%)	42 (22.8%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	9 (9.9%)	2 (2.2%)	11 (6.0%)
	Other evidence of intracranial surgery	15 (16.5%)	12 (12.9%)	27 (14.7%)
	Metallic artifact overlapping region of interest	1 (1.1%)	2 (2.2%)	3 (1.6%)
	Motion artifact overlapping region of interest	4 (4.4%)	2 (2.2%)	6 (3.3%)
	Any finding	89 (97.8%)	52 (55.9%)	141 (76.6%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Mass effect	Slice thickness range: $\leq 1.5\text{mm}$	N=493	N=133	N=626
	Intracranial hemorrhage	320 (64.9%)	6 (4.5%)	326 (52.1%)
	Fracture including calvarial/ skull base/facial	10 (2.0%)	1 (0.8%)	11 (1.8%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	317 (64.3%)	32 (24.1%)	349 (55.8%)
	Parenchymal atrophy (excluding age-expected atrophy)	8 (1.6%)	4 (3.0%)	12 (1.9%)
	Edema including transependymal/vasogenic	318 (64.5%)	5 (3.8%)	323 (51.6%)
	Hydrocephalus including obstructive/non-obstructive	79 (16.0%)	5 (3.8%)	84 (13.4%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	18 (3.7%)	1 (0.8%)	19 (3.0%)
	Other evidence of intracranial surgery	92 (18.7%)	3 (2.3%)	95 (15.2%)
	Metallic artifact overlapping region of interest	28 (5.7%)	7 (5.3%)	35 (5.6%)
	Motion artifact overlapping region of interest	45 (9.1%)	5 (3.8%)	50 (8.0%)
	Any finding	476 (96.6%)	56 (42.1%)	532 (85.0%)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$	N=594	N=168	N=762
	Intracranial hemorrhage	366 (61.6%)	10 (6.0%)	376 (49.3%)
	Fracture including calvarial/ skull base/facial	12 (2.0%)	1 (0.6%)	13 (1.7%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	389 (65.5%)	39 (23.2%)	428 (56.2%)
	Parenchymal atrophy (excluding age-expected atrophy)	8 (1.3%)	5 (3.0%)	13 (1.7%)
	Edema including transependymal/vasogenic	391 (65.8%)	8 (4.8%)	399 (52.4%)
	Hydrocephalus including obstructive/non-obstructive	96 (16.2%)	8 (4.8%)	104 (13.6%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	21 (3.5%)	2 (1.2%)	23 (3.0%)
	Other evidence of intracranial surgery	110 (18.5%)	4 (2.4%)	114 (15.0%)
	Metallic artifact overlapping region of interest	26 (4.4%)	8 (4.8%)	34 (4.5%)
	Motion artifact overlapping region of interest	46 (7.7%)	8 (4.8%)	54 (7.1%)
	Any finding	572 (96.3%)	69 (41.1%)	641 (84.1%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Obstructive hydrocephalus	Slice thickness range: $\leq 1.5\text{mm}$	N=75	N=100	N=175
	Intracranial hemorrhage	51 (68.0%)	32 (32.0%)	83 (47.4%)
	Fracture including calvarial/ skull base/facial	0 (0.0%)	2 (2.0%)	2 (1.1%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	70 (93.3%)	40 (40.0%)	110 (62.9%)
	Parenchymal atrophy (excluding age-expected atrophy)	0 (0.0%)	2 (2.0%)	2 (1.1%)
	Edema including transependymal/vasogenic	71 (94.7%)	22 (22.0%)	93 (53.1%)
	Non-obstructive hydrocephalus	0 (0.0%)	5 (5.0%)	5 (2.9%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	13 (17.3%)	4 (4.0%)	17 (9.7%)
	Other evidence of intracranial surgery	23 (30.7%)	12 (12.0%)	35 (20.0%)
	Metallic artifact overlapping region of interest	2 (2.7%)	3 (3.0%)	5 (2.9%)
	Motion artifact overlapping region of interest	2 (2.7%)	2 (2.0%)	4 (2.3%)
	Any finding	75 (100.0%)	58 (58.0%)	133 (76.0%)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$	N=84	N=107	N=191
	Intracranial hemorrhage	58 (69.0%)	33 (30.8%)	91 (47.6%)
	Fracture including calvarial/ skull base/facial	0 (0.0%)	2 (1.9%)	2 (1.0%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	78 (92.9%)	40 (37.4%)	118 (61.8%)
	Parenchymal atrophy (excluding age-expected atrophy)	0 (0.0%)	2 (1.9%)	2 (1.0%)
	Edema including transependymal/vasogenic	79 (94.0%)	21 (19.6%)	100 (52.4%)
	Non-obstructive hydrocephalus	0 (0.0%)	5 (4.7%)	5 (2.6%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	14 (16.7%)	4 (3.7%)	18 (9.4%)
	Other evidence of intracranial surgery	23 (27.4%)	11 (10.3%)	34 (17.8%)
	Metallic artifact overlapping region of interest	2 (2.4%)	3 (2.8%)	5 (2.6%)
	Motion artifact overlapping region of interest	4 (4.8%)	3 (2.8%)	7 (3.7%)
	Any finding	83 (98.8%)	60 (56.1%)	143 (74.9%)

Finding	Subgroup	Ground truth positive	Ground truth negative	Total
Vasogenic edema	Slice thickness range: $\leq 1.5\text{mm}$	N=60	N=107	N=167
	Intracranial hemorrhage	43 (71.7%)	10 (9.3%)	53 (31.7%)
	Fracture including calvarial/ skull base/facial	1 (1.7%)	1 (0.9%)	2 (1.2%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	56 (93.3%)	27 (25.2%)	83 (49.7%)
	Parenchymal atrophy (excluding age-expected atrophy)	0 (0.0%)	3 (2.8%)	3 (1.8%)
	Edema including transependymal	5 (8.3%)	8 (7.5%)	13 (7.8%)
	Hydrocephalus including obstructive/non-obstructive	13 (21.7%)	9 (8.4%)	22 (13.2%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	7 (11.7%)	4 (3.7%)	11 (6.6%)
	Other evidence of intracranial surgery	18 (30.0%)	5 (4.7%)	23 (13.8%)
	Metallic artifact overlapping region of interest	2 (3.3%)	0 (0.0%)	2 (1.2%)
	Motion artifact overlapping region of interest	6 (10.0%)	8 (7.5%)	14 (8.4%)
	Any finding	59 (98.3%)	50 (46.7%)	109 (65.3%)
	Slice thickness range: $> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$	N=99	N=194	N=293
	Intracranial hemorrhage	59 (59.6%)	14 (7.2%)	73 (24.9%)
	Fracture including calvarial/ skull base/facial	1 (1.0%)	2 (1.0%)	3 (1.0%)
	Parenchymal abnormality including ischemia/ mass/cyst/encephalomalacia	93 (93.9%)	36 (18.6%)	129 (44.0%)
	Parenchymal atrophy (excluding age-expected atrophy)	0 (0.0%)	4 (2.1%)	4 (1.4%)
	Edema including transependymal	8 (8.1%)	11 (5.7%)	19 (6.5%)
	Hydrocephalus including obstructive/non-obstructive	23 (23.2%)	12 (6.2%)	35 (11.9%)
	Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	7 (7.1%)	5 (2.6%)	12 (4.1%)
	Other evidence of intracranial surgery	21 (21.2%)	6 (3.1%)	27 (9.2%)
	Metallic artifact overlapping region of interest	2 (2.0%)	0 (0.0%)	2 (0.7%)
	Motion artifact overlapping region of interest	6 (6.1%)	11 (5.7%)	17 (5.8%)
	Any finding	97 (98.0%)	67 (34.5%)	164 (56.0%)

CTB: Presence of specific co-existing findings or abnormalities for intra-axial hemorrhage

Finding	Operating point	Co-existing findings or anomalies	Absent/ present	Sensitivity (95% CI)	Specificity (95% CI)
Intra-axial hemorrhage	Slice thickness range: $\leq 1.5\text{mm}$				
	0.322700	Other intracranial hemorrhage	Absent	87.6 (82.2,92.3)	97.7 (95.4,99.4)
			Present	96.1 (93.9,98.1)	42.9 (28.6,57.1)*
		Fracture including calvarial/ skull base/facial	Absent	92.7 (90.2,95.0)	86.3 (81.5,91.0)
			Present	97.5 (92.5,100.0)	72.7 (45.5,100.0)
		Parenchymal abnormality including ischemia/mass/ cyst/encephalomalacia	Absent	94.6 (92.1,96.9)	85.6 (79.6,90.4)
			Present	89.0 (83.5,94.5)	85.5 (74.5,94.5)
		Parenchymal atrophy (excluding age-expected atrophy)	Absent	93.6 (91.0,96.0)	85.2 (79.6,90.1)
			Present	91.3 (85.4,96.1)	86.7 (78.3,95.0)
		Edema including transependymal/vasogenic	Absent	71.1 (57.8,84.4)	89.0 (85.0,93.0)
			Present	95.4 (93.3,97.2)	54.5 (31.8,72.7)
		Hydrocephalus including obstructive/non-obstructive	Absent	93.0 (90.5,95.2)	87.2 (82.5,91.5)
			Present	93.9 (89.0,98.8)	54.5 (27.3,81.8)
		Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	Absent	93.2 (90.7,95.4)	85.3 (80.2,89.9)
			Present	92.3 (80.8,100.0)	100.0 (100.0,100.0)
		Other evidence of intracranial surgery	Absent	92.7 (90.0,95.1)	85.9 (81.1,90.3)
			Present	95.8 (90.1,100.0)	81.2 (62.5,100.0)
		Metallic artifact overlapping region of interest	Absent	93.4 (90.9,95.7)	85.8 (80.9,90.2)
			Present	89.7 (79.5,97.4)	83.3 (66.7,100.0)
		Motion artifact overlapping region of interest	Absent	93.5 (91.2,95.9)	85.6 (80.8,89.9)
			Present	89.1 (80.4,97.8)	85.7 (64.3,100.0)
	*Specificity may be reduced for intra-axial hemorrhage in the presence of other intracranial hemorrhage findings.				

Finding	Operating point	Co-existing findings or anomalies	Absent/ present	Sensitivity (95% CI)	Specificity (95% CI)
Intra-axial hemorrhage	Slice thickness range: <1.5mm & ≤5.0mm				
	0.203600	Other intracranial hemorrhage	Absent	88.0 (83.4,92.2)	98.2 (96.3,99.5)
			Present	96.5 (94.6,98.4)	45.1 (33.8,56.3)*
		Fracture including calvarial/ skull base/facial	Absent	93.1 (90.9,95.2)	86.4 (82.0,90.4)
			Present	95.7 (89.4,100.0)	62.5 (37.5,87.5)
		Parenchymal abnormality including ischemia/mass/ cyst/encephalomalacia	Absent	94.5 (92.2,96.3)	87.8 (83.3,91.9)
			Present	90.2 (85.6,94.8)	76.1 (65.7,85.1)
		Parenchymal atrophy (excluding age-expected atrophy)	Absent	93.8 (91.6,96.0)	83.9 (79.0,88.8)
			Present	91.8 (87.3,96.3)	88.0 (80.7,94.0)
		Edema including transependymal/vasogenic	Absent	77.4 (66.1,87.1)	89.0 (85.2,92.8)
			Present	95.2 (93.3,97.0)	41.7 (20.8,62.5)
		Hydrocephalus including obstructive/non-obstructive	Absent	92.9 (90.6,95.1)	85.8 (81.7,89.8)
			Present	95.8 (91.7,99.0)	71.4 (50.0,92.9)
		Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	Absent	93.4 (91.4,95.5)	84.8 (80.5,88.7)
			Present	93.1 (82.8,100.0)	100.0 (100.0,100.0)
		Other evidence of intracranial surgery	Absent	93.0 (90.6,94.9)	85.6 (81.2,89.7)
			Present	95.9 (90.5,100.0)	76.5 (52.9,94.1)
		Metallic artifact overlapping region of interest	Absent	93.6 (91.6,95.6)	85.3 (80.9,89.7)
			Present	90.2 (80.5,97.6)	81.2 (62.5,100.0)
		Motion artifact overlapping region of interest	Absent	93.0 (90.8,95.1)	84.3 (79.9,88.3)
			Present	96.4 (90.9,100.0)	100.0 (100.0,100.0)
		*Specificity may be reduced for intra-axial hemorrhage in the presence of other intracranial hemorrhage findings.			

Finding	Operating point	Co-existing findings or anomalies	Absent/ present	Sensitivity (95% CI)	Specificity (95% CI)
Intra-axial hemorrhage	Slice thickness range: <1.5mm & ≤5.0mm				
	0.322700	Other intracranial hemorrhage	Absent	83.4 (78.3,88.5)	99.1 (97.7,100.0)
			Present	94.3 (91.9,96.5)	63.4 (52.1,74.6)*
		Fracture including calvarial/ skull base/facial	Absent	89.8 (87.2,92.2)	90.8 (87.1,94.1)
			Present	95.7 (89.4,100.0)	81.2 (62.5,100.0)
		Parenchymal abnormality including ischemia/mass/ cyst/encephalomalacia	Absent	91.9 (89.4,94.2)	92.3 (88.7,95.5)
			Present	85.6 (79.7,90.8)	83.6 (74.6,92.5)
		Parenchymal atrophy (excluding age-expected atrophy)	Absent	91.2 (88.5,93.6)	88.8 (84.4,93.2)
			Present	87.3 (81.3,92.5)	94.0 (88.0,98.8)
		Edema including transependymal/vasogenic	Absent	64.5 (53.2,75.8)	93.9 (90.9,96.6)
			Present	93.3 (91.0,95.4)	50.0 (29.2,70.8)
		Hydrocephalus including obstructive/non-obstructive	Absent	89.6 (86.8,92.3)	91.2 (87.6,94.2)
			Present	93.8 (88.5,97.9)	71.4 (50.0,92.9)
		Ventricular drain including extraventricular drain, ventriculoperitoneal shunt	Absent	90.7 (88.4,93.2)	90.1 (86.5,93.3)
			Present	82.8 (69.0,96.6)	100.0 (100.0,100.0)
		Other evidence of intracranial surgery	Absent	89.7 (87.1,92.2)	91.1 (87.8,94.1)
			Present	94.6 (89.2,98.6)	76.5 (52.9,94.1)
		Metallic artifact overlapping region of interest	Absent	90.7 (88.1,93.0)	89.7 (85.7,93.4)
			Present	85.4 (75.5,95.1)	100.0 (100.0,100.0)
		Motion artifact overlapping region of interest	Absent	90.0 (87.6,92.5)	89.8 (85.8,93.4)
			Present	92.7 (85.5,98.2)	100.0 (100.0,100.0)
		*Specificity may be reduced for intra-axial hemorrhage in the presence of other intracranial hemorrhage findings.			

Other

Results

Other subgroup analysis was performed including:

- the sensitivity for each finding by operating point and acute subdural/epidural hematoma subgroup location
- the sensitivity and specificity by operating point for the presence of mimics for obstructive hydrocephalus*

*Parenchymal atrophy (excluding age-expected atrophy), non-obstructive hydrocephalus, age ≥ 75 years and ≥ 80 years may mimic the appearance of obstructive hydrocephalus due to enlargement of ventricles.

See:

- [CTB: Acute subdural/epidural hematoma location](#), below
- [CTB: Presence of mimics for obstructive hydrocephalus - sensitivity and specificity](#), page 68

CTB: Acute subdural/epidural hematoma location

Finding	Slice thickness range	Operating point	Subgroup	Sensitivity % (95% CI)
Acute subdural/epidural hematoma	$\leq 1.5\text{mm}$	0.060177	Epidural	94.7 (82.4,100.0)
			Subdural	90.5 (86.7,94.0)
			Subdural and epidural	96.8 (89.7,100.0)
		0.101143	Epidural	94.7 (82.4,100.0)
			Subdural	87.7 (83.7,91.6)
			Subdural and epidural	96.8 (89.7,100.0)
		0.135700	Epidural	84.2 (66.6,100.0)
			Subdural	86.2 (81.9,90.2)
			Subdural and epidural	90.3 (79.2,100.0)
	$> 1.5\text{mm} \ \& \ \leq 5.0\text{mm}$	0.060177	Epidural	80.0 (64.0,95.5)
			Subdural	80.8 (76.6,85.0)
			Subdural and epidural	97.4 (91.7,100.0)

CTB: Presence of mimics for obstructive hydrocephalus - sensitivity and specificity

Finding	Operating point	Co-existing findings or anomalies	Absent/ present	n	Sensitivity (95% CI)	Specificity (95% CI)
Obstructive hydrocephalus	Slice thickness range: $\leq 1.5\text{mm}$					
	0.149943	Parenchymal atrophy (excluding age-expected atrophy)	Absent	173	97.3 (93.3,100.0)	93.9 (88.8,98.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	170	97.3 (93.3,100.0)	94.7 (89.5,98.9)
			Present	5	-	80.0 (40.0,100.0)
		Age ≥ 75 years	Absent	129	96.6 (91.4,100.0)	93.0 (87.3,98.6)
			Present	46	100.0 (100.0,100.0)	96.6 (89.7,100.0)
		Age ≥ 80 years	Absent	150	96.9 (92.3,100.0)	92.9 (87.1,97.6)
			Present	25	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	146	96.6 (91.4,100.0)	94.0 (88.1,98.5)
			Present	29	100.0 (100.0,100.0)	93.9 (84.8,100.0)
	0.185900	Parenchymal atrophy (excluding age-expected atrophy)	Absent	173	94.7 (89.3,98.7)	94.9 (90.8,99.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	170	94.7 (89.3,98.7)	95.8 (91.6,98.9)
			Present	5	-	80.0 (40.0,100.0)
		Age ≥ 75 years	Absent	129	93.1 (86.2,98.3)	94.4 (88.7,98.6)
			Present	46	100.0 (100.0,100.0)	96.6 (89.7,100.0)
		Age ≥ 80 years	Absent	150	93.8 (87.7,98.5)	94.1 (89.4,98.8)
			Present	25	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	146	93.1 (86.2,98.3)	95.5 (89.6,100.0)
			Present	29	100.0 (100.0,100.0)	93.9 (84.8,100.0)
	0.281473	Parenchymal atrophy (excluding age-expected atrophy)	Absent	173	92.0 (85.3,97.3)	96.9 (93.9,100.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	170	92.0 (85.3,97.3)	97.9 (94.7,100.0)
			Present	5	-	80.0 (40.0,100.0)
		Age ≥ 75 years	Absent	129	91.4 (84.5,98.3)	97.2 (93.0,100.0)
			Present	46	94.1 (82.4,100.0)	96.6 (89.7,100.0)
		Age ≥ 80 years	Absent	150	90.8 (83.1,96.9)	96.5 (91.8,100.0)
			Present	25	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	146	91.4 (82.8,98.3)	98.5 (95.5,100.0)
			Present	29	94.1 (82.4,100.0)	93.9 (84.8,100.0)

Finding	Operating point	Co-existing findings or anomalies	Absent/present	n	Sensitivity (95% CI)	Specificity (95% CI)
Obstructive hydrocephalus	Slice thickness range: <1.5mm & ≤5.0mm					
	0.100591	Parenchymal atrophy (excluding age-expected atrophy)	Absent	189	97.6 (94.0,100.0)	95.2 (90.5,99.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	186	97.6 (94.0,100.0)	96.1 (92.2,99.0)
			Present	5	-	80.0 (33.3,100.0)
		Age ≥ 75 years	Absent	140	96.9 (92.2,100.0)	94.7 (89.5,98.7)
			Present	51	100.0 (100.0,100.0)	96.8 (90.3,100.0)
		Age ≥ 80 years	Absent	163	97.2 (93.1,100.0)	94.5 (89.0,98.9)
			Present	28	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	136	96.9 (92.2,100.0)	95.8 (90.3,100.0)
			Present	55	100.0 (100.0,100.0)	94.3 (85.7,100.0)
	0.149943	Parenchymal atrophy (excluding age-expected atrophy)	Absent	189	95.2 (90.5,98.8)	95.2 (90.5,99.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	186	95.2 (90.5,98.8)	96.1 (92.2,99.0)
			Present	5	-	80.0 (33.3,100.0)
		Age ≥ 75 years	Absent	140	93.8 (87.5,98.4)	94.7 (89.5,98.7)
			Present	51	100.0 (100.0,100.0)	96.8 (90.3,100.0)
		Age ≥ 80 years	Absent	163	94.4 (88.9,98.6)	94.5 (89.0,98.9)
			Present	28	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	136	93.8 (87.5,98.4)	95.8 (90.3,100.0)
			Present	55	100.0 (100.0,100.0)	94.3 (85.7,100.0)
	0.185900	Parenchymal atrophy (excluding age-expected atrophy)	Absent	189	94.0 (88.1,98.8)	95.2 (90.5,99.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	186	94.0 (88.1,98.8)	96.1 (92.2,99.0)
			Present	5	-	80.0 (33.3,100.0)
		Age ≥ 75 years	Absent	140	92.2 (84.4,98.4)	94.7 (89.5,98.7)
			Present	51	100.0 (100.0,100.0)	96.8 (90.3,100.0)
		Age ≥ 80 years	Absent	163	93.1 (87.5,98.6)	94.5 (89.0,98.9)
			Present	28	100.0 (100.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	136	92.2 (84.4,98.4)	95.8 (90.3,100.0)
			Present	55	100.0 (100.0,100.0)	94.3 (85.7,100.0)

Finding	Operating point	Co-existing findings or anomalies	Absent/ present	n	Sensitivity (95% CI)	Specificity (95% CI)
Obstructive hydrocephalus	0.281473	Parenchymal atrophy (excluding age-expected atrophy)	Absent	189	88.1 (81.0,94.1)	95.2 (90.5,99.0)
			Present	2	-	100.0 (100.0,100.0)
		Non-obstructive hydrocephalus	Absent	186	88.1 (81.0,94.1)	96.1 (92.2,99.0)
			Present	5	-	80.0 (33.3,100.0)
		Age ≥ 75 years	Absent	140	89.1 (81.2,95.3)	94.7 (89.5,98.7)
			Present	51	85.0 (70.0,100.0)	96.8 (90.3,100.0)
		Age ≥ 80 years	Absent	163	87.5 (79.2,94.4)	94.5 (89.0,98.9)
			Present	28	91.7 (75.0,100.0)	100.0 (100.0,100.0)
		Any mimic	Absent	136	89.1 (81.2,95.3)	95.8 (90.3,100.0)
			Present	55	85.0 (70.0,100.0)	94.3 (85.7,100.0)

Triage effectiveness performance

Triage effectiveness

Definition

Triage turnaround time is defined as the time taken for the device to process and package the study, perform model inference, process inference results, package them as a triage payload and transmit the triage payload to the relevant customer platform.

Assessing triage effectiveness

Triage turnaround time of the device was assessed using validation datasets of cases positive for each finding eligible for prioritization, as shown in the tables below.

These cases were collected from multiple data sources spanning a variety of geographical locations, patient demographics and technical characteristics.

Results

CXR: Algorithm speed

The findings marked with an asterisk (*) were calculated together.

Finding	No. of cases	Mean (sec)	Std deviation
Pleural effusion	n = 303	42.3	10.0
Pneumoperitoneum			
Pneumothorax*			
Tension pneumothorax*			
Vertebral compression fracture			

CTB: Algorithm speed

The following findings were calculated together.

Finding	No. of cases	Mean (sec)	Std deviation
Acute subarachnoid hemorrhage	n = 277	81.6	10.8
Acute subdural/epidural hematoma			
Intra-axial hemorrhage			
Intraventricular hemorrhage)			
Mass effect			
Obstructive hydrocephalus			
Vasogenic edema			

Support

Support

Support and feedback

Refer to the following table for support and feedback details:

Support type	Details
Professional services, technical support, product feedback and complaints	Email support@annalise.ai Any serious incidents related to the device/s should be reported to Annalise.ai, and where applicable, the competent authority or regulatory authority in which the user and/or patient is established.
Product user, performance and administration guides	Check our website: annalise.ai/guides

△ annalise